

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: O3631

Client: LaBella Associates P.C.

Analytical Method: SW8260-Low

Datafile : VN078511.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VN0714WBS01	Dichlorodifluoromethane	20	18.6	ug/L	93			69	116	
	Chloromethane	20	21.7	ug/L	109			65	116	
	Vinyl chloride	20	19.5	ug/L	98			65	117	
	Bromomethane	20	18.7	ug/L	94			42	173	
	Chloroethane	20	20.8	ug/L	104			44	165	
	Trichlorofluoromethane	20	22.8	ug/L	114			73	115	
	1,1,2-Trichlorotrifluoroethane	20	19.8	ug/L	99			80	112	
	1,1-Dichloroethene	20	19.6	ug/L	98			74	110	
	Acetone	100	110	ug/L	110			60	125	
	Carbon disulfide	20	17.6	ug/L	88			64	112	
	Methyl tert-butyl Ether	20	20.2	ug/L	101			78	114	
	Methyl Acetate	20	20.2	ug/L	101			67	125	
	Methylene Chloride	20	20.6	ug/L	103			72	114	
	trans-1,2-Dichloroethene	20	19.7	ug/L	99			75	108	
	1,1-Dichloroethane	20	20.0	ug/L	100			78	112	
	Cyclohexane	20	19.1	ug/L	96			75	110	
	2-Butanone	100	110	ug/L	110			65	122	
	Carbon Tetrachloride	20	20.3	ug/L	102			77	113	
	cis-1,2-Dichloroethene	20	19.8	ug/L	99			77	110	
	Bromochloromethane	20	19.7	ug/L	99			70	124	
	Chloroform	20	21.3	ug/L	106			79	113	
	1,1,1-Trichloroethane	20	20.3	ug/L	102			80	108	
	Methylcyclohexane	20	19.7	ug/L	99			72	115	
	Benzene	20	20.6	ug/L	103			82	109	
	1,2-Dichloroethane	20	20.4	ug/L	102			80	115	
	Trichloroethene	20	19.6	ug/L	98			77	113	
	1,2-Dichloropropane	20	20.5	ug/L	103			83	111	
	Bromodichloromethane	20	20.6	ug/L	103			83	110	
	4-Methyl-2-Pentanone	100	110	ug/L	110			74	118	
	Toluene	20	20.3	ug/L	102			82	110	
	t-1,3-Dichloropropene	20	20.5	ug/L	103			79	110	
	cis-1,3-Dichloropropene	20	21.0	ug/L	105			82	110	
	1,1,2-Trichloroethane	20	21.3	ug/L	106			83	112	
	2-Hexanone	100	110	ug/L	110			73	117	
	Dibromochloromethane	20	21.5	ug/L	108			82	110	
	1,2-Dibromoethane	20	20.6	ug/L	103			81	110	
	Tetrachloroethene	20	20.3	ug/L	102			67	123	
	Chlorobenzene	20	19.7	ug/L	99			82	109	
	Ethyl Benzene	20	20.0	ug/L	100			83	109	
	m/p-Xylenes	40	40.9	ug/L	102			82	110	
	o-Xylene	20	19.8	ug/L	99			83	109	
	Styrene	20	20.0	ug/L	100			80	111	
	Bromoform	20	20.6	ug/L	103			79	109	
	Isopropylbenzene	20	19.0	ug/L	95			83	112	
	1,1,2,2-Tetrachloroethane	20	21.4	ug/L	107			76	118	
	N-propylbenzene	20	19.6	ug/L	98			83	112	
	1,3,5-Trimethylbenzene	20	19.8	ug/L	99			85	112	
	tert-Butylbenzene	20	19.3	ug/L	97			83	112	
	1,2,4-Trimethylbenzene	20	20.0	ug/L	100			85	111	



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: O3631

Client: LaBella Associates P.C.

Analytical Method: SW8260-Low

Datafile : VN078511.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VN0714WBS01	Sec-butylbenzene	20	19.4	ug/L	97			81	114	
	p-Isopropyltoluene	20	19.6	ug/L	98			78	116	
	1,3-Dichlorobenzene	20	19.6	ug/L	98			82	108	
	1,4-Dichlorobenzene	20	18.2	ug/L	91			82	107	
	n-Butylbenzene	20	19.8	ug/L	99			75	115	
	1,2-Dichlorobenzene	20	18.7	ug/L	94			82	109	
	1,2-Dibromo-3-Chloropropane	20	20.3	ug/L	102			68	112	
	1,2,4-Trichlorobenzene	20	16.4	ug/L	82			74	114	
	1,2,3-Trichlorobenzene	20	16.0	ug/L	80			77	113	