



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: O3631

Client: LaBella Associates P.C.

Analytical Method: SW8260-Low

Datafile : VN078512.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VN0714WBSD01	Dichlorodifluoromethane	20	17.9	ug/L	90	3		69	116	20
	Chloromethane	20	20.9	ug/L	104	5		65	116	20
	Vinyl chloride	20	19.0	ug/L	95	3		65	117	20
	Bromomethane	20	19.0	ug/L	95	1		42	173	20
	Chloroethane	20	19.0	ug/L	95	9		44	165	20
	Trichlorofluoromethane	20	19.2	ug/L	96	17		73	115	20
	1,1,2-Trichlorotrifluoroethane	20	20.6	ug/L	103	4		80	112	20
	1,1-Dichloroethene	20	19.1	ug/L	96	2		74	110	20
	Acetone	100	100	ug/L	100	10		60	125	20
	Carbon disulfide	20	17.0	ug/L	85	3		64	112	20
	Methyl tert-butyl Ether	20	20.9	ug/L	104	3		78	114	20
	Methyl Acetate	20	20.5	ug/L	103	2		67	125	20
	Methylene Chloride	20	21.0	ug/L	105	2		72	114	20
	trans-1,2-Dichloroethene	20	19.4	ug/L	97	2		75	108	20
	1,1-Dichloroethane	20	20.2	ug/L	101	1		78	112	20
	Cyclohexane	20	18.1	ug/L	91	5		75	110	20
	2-Butanone	100	110	ug/L	110	0		65	122	20
	Carbon Tetrachloride	20	20.1	ug/L	101	1		77	113	20
	cis-1,2-Dichloroethene	20	19.9	ug/L	100	1		77	110	20
	Bromochloromethane	20	19.7	ug/L	99	0		70	124	20
	Chloroform	20	21.0	ug/L	105	1		79	113	20
	1,1,1-Trichloroethane	20	20.4	ug/L	102	0		80	108	20
	Methylcyclohexane	20	19.5	ug/L	98	1		72	115	20
	Benzene	20	20.6	ug/L	103	0		82	109	20
	1,2-Dichloroethane	20	21.4	ug/L	107	5		80	115	20
	Trichloroethene	20	18.5	ug/L	93	5		77	113	20
	1,2-Dichloropropane	20	20.6	ug/L	103	0		83	111	20
	Bromodichloromethane	20	21.4	ug/L	107	4		83	110	20
	4-Methyl-2-Pentanone	100	110	ug/L	110	0		74	118	20
	Toluene	20	20.3	ug/L	102	0		82	110	20
	t-1,3-Dichloropropene	20	20.7	ug/L	104	1		79	110	20
	cis-1,3-Dichloropropene	20	21.1	ug/L	106	1		82	110	20
	1,1,2-Trichloroethane	20	21.9	ug/L	110	4		83	112	20
	2-Hexanone	100	110	ug/L	110	0		73	117	20
	Dibromochloromethane	20	21.5	ug/L	108	0		82	110	20
	1,2-Dibromoethane	20	21.1	ug/L	106	3		81	110	20
	Tetrachloroethene	20	20.1	ug/L	101	1		67	123	20
	Chlorobenzene	20	19.4	ug/L	97	2		82	109	20
	Ethyl Benzene	20	19.8	ug/L	99	1		83	109	20
	m/p-Xylenes	40	40.8	ug/L	102	0		82	110	20
	o-Xylene	20	19.6	ug/L	98	1		83	109	20
	Styrene	20	20.1	ug/L	101	1		80	111	20
	Bromoform	20	20.6	ug/L	103	0		79	109	20
	Isopropylbenzene	20	18.3	ug/L	92	3		83	112	20
	1,1,2,2-Tetrachloroethane	20	21.2	ug/L	106	1		76	118	20
	N-propylbenzene	20	18.8	ug/L	94	4		83	112	20
	1,3,5-Trimethylbenzene	20	19.0	ug/L	95	4		85	112	20
	tert-Butylbenzene	20	18.5	ug/L	93	4		83	112	20
	1,2,4-Trimethylbenzene	20	19.6	ug/L	98	2		85	111	20



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: O3631

Client: LaBella Associates P.C.

Analytical Method: SW8260-Low

Datafile : VN078512.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VN0714WBSD01	Sec-butylbenzene	20	18.8	ug/L	94	3		81	114	20
	p-Isopropyltoluene	20	19.0	ug/L	95	3		78	116	20
	1,3-Dichlorobenzene	20	19.0	ug/L	95	3		82	108	20
	1,4-Dichlorobenzene	20	18.0	ug/L	90	1		82	107	20
	n-Butylbenzene	20	19.2	ug/L	96	3		75	115	20
	1,2-Dichlorobenzene	20	19.2	ug/L	96	2		82	109	20
	1,2-Dibromo-3-Chloropropane	20	19.7	ug/L	99	3		68	112	20
	1,2,4-Trichlorobenzene	20	16.1	ug/L	81	1		74	114	20
	1,2,3-Trichlorobenzene	20	16.7	ug/L	84	5		77	113	20