



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: O3631

Client: LaBella Associates P.C.

Analytical Method: SW8260D

Datafile : VY014665.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VY0717SBS01	Dichlorodifluoromethane	20	23.5	ug/Kg	117			64	136	
	Chloromethane	20	25.4	ug/Kg	127			70	130	
	Vinyl chloride	20	24.9	ug/Kg	125			72	129	
	Bromomethane	20	25.0	ug/Kg	125			58	141	
	Chloroethane	20	24.9	ug/Kg	125			69	130	
	Trichlorofluoromethane	20	22.2	ug/Kg	111			69	134	
	1,1,2-Trichlorotrifluoroethane	20	21.7	ug/Kg	109			81	123	
	1,1-Dichloroethene	20	22.1	ug/Kg	111			79	121	
	Acetone	100	120	ug/Kg	120			60	131	
	Carbon disulfide	20	22.2	ug/Kg	111			45	154	
	Methyl tert-butyl Ether	20	23.4	ug/Kg	117			77	129	
	Methyl Acetate	20	25.4	ug/Kg	127			69	149	
	Methylene Chloride	20	23.3	ug/Kg	117			39	175	
	trans-1,2-Dichloroethene	20	22.4	ug/Kg	112			80	123	
	1,1-Dichloroethane	20	22.1	ug/Kg	111			82	123	
	Cyclohexane	20	20.5	ug/Kg	103			76	122	
	2-Butanone	100	130	ug/Kg	130			69	131	
	Carbon Tetrachloride	20	21.5	ug/Kg	108			76	129	
	cis-1,2-Dichloroethene	20	22.9	ug/Kg	115			82	123	
	Bromochloromethane	20	23.6	ug/Kg	118			62	134	
	Chloroform	20	22.4	ug/Kg	112			82	125	
	1,1,1-Trichloroethane	20	21.7	ug/Kg	109			80	126	
	Methylcyclohexane	20	20.7	ug/Kg	104			77	123	
	Benzene	20	22.2	ug/Kg	111			84	121	
	1,2-Dichloroethane	20	22.5	ug/Kg	113			81	126	
	Trichloroethene	20	22.2	ug/Kg	111			83	122	
	1,2-Dichloropropane	20	21.8	ug/Kg	109			83	122	
	Bromodichloromethane	20	21.8	ug/Kg	109			82	123	
	4-Methyl-2-Pentanone	100	120	ug/Kg	120			70	135	
	Toluene	20	22.1	ug/Kg	111			83	122	
	t-1,3-Dichloropropene	20	22.1	ug/Kg	111			78	124	
	cis-1,3-Dichloropropene	20	21.8	ug/Kg	109			81	122	
	1,1,2-Trichloroethane	20	22.9	ug/Kg	115			82	125	
	2-Hexanone	100	120	ug/Kg	120			66	138	
	Dibromochloromethane	20	22.8	ug/Kg	114			79	125	
	1,2-Dibromoethane	20	23.4	ug/Kg	117			80	125	
	Tetrachloroethene	20	22.9	ug/Kg	115			83	125	
	Chlorobenzene	20	22.0	ug/Kg	110			84	122	
	Ethyl Benzene	20	22.1	ug/Kg	111			82	124	
	m/p-Xylenes	40	43.0	ug/Kg	108			83	124	
	o-Xylene	20	22.0	ug/Kg	110			83	123	
	Styrene	20	22.2	ug/Kg	111			82	124	
	Bromoform	20	23.6	ug/Kg	118			75	127	
	Isopropylbenzene	20	21.0	ug/Kg	105			82	124	
	1,1,2,2-Tetrachloroethane	20	22.5	ug/Kg	113			77	127	
	N-propylbenzene	20	20.1	ug/Kg	101			81	123	
	1,3,5-Trimethylbenzene	20	20.9	ug/Kg	104			82	124	
	tert-Butylbenzene	20	20.3	ug/Kg	102			81	125	
	1,2,4-Trimethylbenzene	20	21.3	ug/Kg	106			81	125	



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: O3631

Client: LaBella Associates P.C.

Analytical Method: SW8260D

Datafile : VY014665.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VY0717SBS01	Sec-butylbenzene	20	20.2	ug/Kg	101			80	124	
	p-Isopropyltoluene	20	20.5	ug/Kg	103			81	125	
	1,3-Dichlorobenzene	20	21.6	ug/Kg	108			83	122	
	1,4-Dichlorobenzene	20	21.5	ug/Kg	108			84	121	
	n-Butylbenzene	20	20.0	ug/Kg	100			78	126	
	1,2-Dichlorobenzene	20	21.7	ug/Kg	109			83	124	
	1,2-Dibromo-3-Chloropropane	20	23.7	ug/Kg	119			66	134	
	1,2,4-Trichlorobenzene	20	22.0	ug/Kg	110			78	127	
	1,2,3-Trichlorobenzene	20	21.6	ug/Kg	108			70	137	