



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LABE01
Lab Code: CHEM Case No.: 03631 SAS No.: 03631 SDG NO.: 03631
Lab File ID: VN078508.D Date Analyzed: 07/14/2023
Instrument ID: MSVOA_N Time Analyzed: 09:12
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	480719	8.23	810635	9.11	725667	11.87
UPPER LIMIT	961438	8.73	1621270	9.606	1451330	12.371
LOWER LIMIT	240360	7.73	405318	8.606	362834	11.371
EPA SAMPLE NO.						
RINSATE-BLANK	401816	8.24	736822	9.11	683725	11.87
VN0714WBL01	410410	8.24	730500	9.11	682507	11.87
VN0714WBS01	448999	8.24	779605	9.11	687589	11.87
VN0714WBSD01	423054	8.23	721606	9.11	652448	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.