



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LABE01
Lab Code: CHEM Case No.: O3631 SAS No.: O3631 SDG NO.: O3631
Lab File ID: VY014663.D Date Analyzed: 07/17/2023
Instrument ID: MSVOA_Y Time Analyzed: 11:00
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	223671	13.422				
UPPER LIMIT	447342	13.922				
LOWER LIMIT	111836	12.922				
EPA SAMPLE NO.						
SB-02 (2-4)	150203	13.42				
SB-02 (2-4)MS	96285 *	13.42				
SB-02 (2-4)MSD	114737	13.42				
SB-04 (2-4)	172137	13.42				
SB-07 (3-5)	169405	13.42				
SB-09 (3-5)	160685	13.42				
SB-10 (5-6)	155439	13.42				
SB-11 (3.5-4.5)	164868	13.42				
DUP	168156	13.42				
VY0717SBL01	180919	13.42				
VY0717SBS01	217159	13.42				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.