

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV75	Mercury	4.24	4.0	106	90 - 110	CV	07/19/2023	09:50	LB126505

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV61	Mercury	4.87	5.0	97	90 - 110	CV	07/19/2023	09:55	LB126505
CCV62	Mercury	4.87	5.0	98	90 - 110	CV	07/19/2023	10:22	LB126505
CCV63	Mercury	4.90	5.0	98	90 - 110	CV	07/19/2023	10:31	LB126505

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV76	Mercury	4.14	4.0	104	90 - 110	CV	07/19/2023	11:49	LB126507

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV64	Mercury	4.84	5.0	97	90 - 110	CV	07/19/2023	11:53	LB126507
CCV65	Mercury	4.86	5.0	97	90 - 110	CV	07/19/2023	12:23	LB126507
CCV66	Mercury	4.88	5.0	98	90 - 110	CV	07/19/2023	12:50	LB126507
CCV67	Mercury	4.91	5.0	98	90 - 110	CV	07/19/2023	13:17	LB126507
CCV68	Mercury	4.96	5.0	99	90 - 110	CV	07/19/2023	13:44	LB126507
CCV69	Mercury	5.03	5.0	101	90 - 110	CV	07/19/2023	14:07	LB126507
CCV70	Mercury	4.93	5.0	99	90 - 110	CV	07/19/2023	14:42	LB126507
CCV71	Mercury	5.05	5.0	101	90 - 110	CV	07/19/2023	14:59	LB126507

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1030	1000	103	90 - 110	P	07/19/2023	13:05	LB126517
	Barium	502	520	96	90 - 110	P	07/19/2023	13:05	LB126517
	Beryllium	486	510	95	90 - 110	P	07/19/2023	13:05	LB126517
	Cadmium	503	510	99	90 - 110	P	07/19/2023	13:05	LB126517
	Chromium	504	520	97	90 - 110	P	07/19/2023	13:05	LB126517
	Copper	499	510	98	90 - 110	P	07/19/2023	13:05	LB126517
	Lead	991	1000	99	90 - 110	P	07/19/2023	13:05	LB126517
	Manganese	506	520	97	90 - 110	P	07/19/2023	13:05	LB126517
	Nickel	490	530	92	90 - 110	P	07/19/2023	13:05	LB126517
	Selenium	1030	1000	103	90 - 110	P	07/19/2023	13:05	LB126517
	Silver	248	250	99	90 - 110	P	07/19/2023	13:05	LB126517
	Zinc	991	1000	99	90 - 110	P	07/19/2023	13:05	LB126517

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5010	5000	100	90 - 110	P	07/19/2023	13:37	LB126517
	Barium	10000	10000	100	90 - 110	P	07/19/2023	13:37	LB126517
	Beryllium	238	250	95	90 - 110	P	07/19/2023	13:37	LB126517
	Cadmium	2430	2500	97	90 - 110	P	07/19/2023	13:37	LB126517
	Chromium	988	1000	99	90 - 110	P	07/19/2023	13:37	LB126517
	Copper	1210	1250	97	90 - 110	P	07/19/2023	13:37	LB126517
	Lead	4910	5000	98	90 - 110	P	07/19/2023	13:37	LB126517
	Manganese	2440	2500	98	90 - 110	P	07/19/2023	13:37	LB126517
	Nickel	2450	2500	98	90 - 110	P	07/19/2023	13:37	LB126517
	Selenium	5000	5000	100	90 - 110	P	07/19/2023	13:37	LB126517
	Silver	1270	1250	102	90 - 110	P	07/19/2023	13:37	LB126517
	Zinc	2500	2500	100	90 - 110	P	07/19/2023	13:37	LB126517

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLCCV01	Arsenic	20.1	20.0	100	80 - 120	P	07/19/2023	13:45	LB126517
	Barium	104	100	104	80 - 120	P	07/19/2023	13:45	LB126517
	Beryllium	6.35	6.0	106	80 - 120	P	07/19/2023	13:45	LB126517
	Cadmium	6.59	6.0	110	80 - 120	P	07/19/2023	13:45	LB126517
	Chromium	10.4	10.0	104	80 - 120	P	07/19/2023	13:45	LB126517
	Copper	23.2	20.0	116	80 - 120	P	07/19/2023	13:45	LB126517
	Lead	11.3	12.0	94	80 - 120	P	07/19/2023	13:45	LB126517
	Manganese	21.3	20.0	106	80 - 120	P	07/19/2023	13:45	LB126517
	Nickel	40.2	40.0	101	80 - 120	P	07/19/2023	13:45	LB126517
	Selenium	20.3	20.0	102	80 - 120	P	07/19/2023	13:45	LB126517
	Silver	11.1	10.0	111	80 - 120	P	07/19/2023	13:45	LB126517
	Zinc	43.2	40.0	108	80 - 120	P	07/19/2023	13:45	LB126517

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Arsenic	5080	5000	102	90 - 110	P	07/19/2023	14:54	LB126517
	Barium	10200	10000	102	90 - 110	P	07/19/2023	14:54	LB126517
	Beryllium	238	250	95	90 - 110	P	07/19/2023	14:54	LB126517
	Cadmium	2430	2500	97	90 - 110	P	07/19/2023	14:54	LB126517
	Chromium	991	1000	99	90 - 110	P	07/19/2023	14:54	LB126517
	Copper	1250	1250	100	90 - 110	P	07/19/2023	14:54	LB126517
	Lead	4930	5000	99	90 - 110	P	07/19/2023	14:54	LB126517
	Manganese	2450	2500	98	90 - 110	P	07/19/2023	14:54	LB126517
	Nickel	2460	2500	98	90 - 110	P	07/19/2023	14:54	LB126517
	Selenium	5040	5000	101	90 - 110	P	07/19/2023	14:54	LB126517
	Silver	1280	1250	102	90 - 110	P	07/19/2023	14:54	LB126517
	Zinc	2510	2500	100	90 - 110	P	07/19/2023	14:54	LB126517
CCV03	Arsenic	5120	5000	102	90 - 110	P	07/19/2023	16:03	LB126517
	Barium	10300	10000	103	90 - 110	P	07/19/2023	16:03	LB126517
	Beryllium	236	250	94	90 - 110	P	07/19/2023	16:03	LB126517
	Cadmium	2400	2500	96	90 - 110	P	07/19/2023	16:03	LB126517
	Chromium	984	1000	98	90 - 110	P	07/19/2023	16:03	LB126517
	Copper	1270	1250	102	90 - 110	P	07/19/2023	16:03	LB126517
	Lead	4910	5000	98	90 - 110	P	07/19/2023	16:03	LB126517
	Manganese	2460	2500	98	90 - 110	P	07/19/2023	16:03	LB126517
	Nickel	2450	2500	98	90 - 110	P	07/19/2023	16:03	LB126517
	Selenium	5090	5000	102	90 - 110	P	07/19/2023	16:03	LB126517
	Silver	1270	1250	102	90 - 110	P	07/19/2023	16:03	LB126517
	Zinc	2510	2500	100	90 - 110	P	07/19/2023	16:03	LB126517
CCV04	Arsenic	5160	5000	103	90 - 110	P	07/19/2023	16:51	LB126517
	Barium	10300	10000	103	90 - 110	P	07/19/2023	16:51	LB126517
	Beryllium	236	250	95	90 - 110	P	07/19/2023	16:51	LB126517
	Cadmium	2420	2500	97	90 - 110	P	07/19/2023	16:51	LB126517
	Chromium	989	1000	99	90 - 110	P	07/19/2023	16:51	LB126517
	Copper	1270	1250	102	90 - 110	P	07/19/2023	16:51	LB126517
	Lead	4930	5000	99	90 - 110	P	07/19/2023	16:51	LB126517
	Manganese	2460	2500	99	90 - 110	P	07/19/2023	16:51	LB126517
	Nickel	2460	2500	98	90 - 110	P	07/19/2023	16:51	LB126517
	Selenium	5140	5000	103	90 - 110	P	07/19/2023	16:51	LB126517

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Silver	1270	1250	101	90 - 110	P	07/19/2023	16:51	LB126517
	Zinc	2520	2500	101	90 - 110	P	07/19/2023	16:51	LB126517
CCV05	Arsenic	5200	5000	104	90 - 110	P	07/19/2023	17:44	LB126517
	Barium	10400	10000	104	90 - 110	P	07/19/2023	17:44	LB126517
	Beryllium	238	250	95	90 - 110	P	07/19/2023	17:44	LB126517
	Cadmium	2420	2500	97	90 - 110	P	07/19/2023	17:44	LB126517
	Chromium	984	1000	98	90 - 110	P	07/19/2023	17:44	LB126517
	Copper	1280	1250	102	90 - 110	P	07/19/2023	17:44	LB126517
	Lead	4950	5000	99	90 - 110	P	07/19/2023	17:44	LB126517
	Manganese	2450	2500	98	90 - 110	P	07/19/2023	17:44	LB126517
	Nickel	2480	2500	99	90 - 110	P	07/19/2023	17:44	LB126517
	Selenium	5180	5000	104	90 - 110	P	07/19/2023	17:44	LB126517
CCV06	Silver	1260	1250	101	90 - 110	P	07/19/2023	17:44	LB126517
	Zinc	2570	2500	103	90 - 110	P	07/19/2023	17:44	LB126517
	Arsenic	5050	5000	101	90 - 110	P	07/19/2023	18:39	LB126517
	Barium	10300	10000	103	90 - 110	P	07/19/2023	18:39	LB126517
	Beryllium	235	250	94	90 - 110	P	07/19/2023	18:39	LB126517
	Cadmium	2360	2500	94	90 - 110	P	07/19/2023	18:39	LB126517
	Chromium	978	1000	98	90 - 110	P	07/19/2023	18:39	LB126517
	Copper	1250	1250	100	90 - 110	P	07/19/2023	18:39	LB126517
	Lead	4830	5000	96	90 - 110	P	07/19/2023	18:39	LB126517
	Manganese	2420	2500	97	90 - 110	P	07/19/2023	18:39	LB126517
CCV07	Nickel	2420	2500	97	90 - 110	P	07/19/2023	18:39	LB126517
	Selenium	5050	5000	101	90 - 110	P	07/19/2023	18:39	LB126517
	Silver	1270	1250	101	90 - 110	P	07/19/2023	18:39	LB126517
	Zinc	2530	2500	101	90 - 110	P	07/19/2023	18:39	LB126517
	Arsenic	5030	5000	101	90 - 110	P	07/19/2023	19:26	LB126517
	Barium	10100	10000	101	90 - 110	P	07/19/2023	19:26	LB126517
	Beryllium	233	250	93	90 - 110	P	07/19/2023	19:26	LB126517
	Cadmium	2360	2500	94	90 - 110	P	07/19/2023	19:26	LB126517
	Chromium	960	1000	96	90 - 110	P	07/19/2023	19:26	LB126517
	Copper	1220	1250	98	90 - 110	P	07/19/2023	19:26	LB126517
	Lead	4840	5000	97	90 - 110	P	07/19/2023	19:26	LB126517
	Manganese	2390	2500	96	90 - 110	P	07/19/2023	19:26	LB126517

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Nickel	2410	2500	97	90 - 110	P	07/19/2023	19:26	LB126517
	Selenium	5030	5000	101	90 - 110	P	07/19/2023	19:26	LB126517
	Silver	1240	1250	100	90 - 110	P	07/19/2023	19:26	LB126517
	Zinc	2480	2500	99	90 - 110	P	07/19/2023	19:26	LB126517
CCV08	Arsenic	5020	5000	100	90 - 110	P	07/19/2023	20:14	LB126517
	Barium	10000	10000	100	90 - 110	P	07/19/2023	20:14	LB126517
	Beryllium	233	250	93	90 - 110	P	07/19/2023	20:14	LB126517
	Cadmium	2380	2500	95	90 - 110	P	07/19/2023	20:14	LB126517
	Chromium	966	1000	97	90 - 110	P	07/19/2023	20:14	LB126517
	Copper	1210	1250	97	90 - 110	P	07/19/2023	20:14	LB126517
	Lead	4840	5000	97	90 - 110	P	07/19/2023	20:14	LB126517
	Manganese	2370	2500	95	90 - 110	P	07/19/2023	20:14	LB126517
	Nickel	2430	2500	97	90 - 110	P	07/19/2023	20:14	LB126517
	Selenium	5030	5000	101	90 - 110	P	07/19/2023	20:14	LB126517
	Silver	1250	1250	100	90 - 110	P	07/19/2023	20:14	LB126517
	Zinc	2500	2500	100	90 - 110	P	07/19/2023	20:14	LB126517
CCV09	Arsenic	5170	5000	103	90 - 110	P	07/19/2023	21:03	LB126517
	Barium	10300	10000	103	90 - 110	P	07/19/2023	21:03	LB126517
	Beryllium	244	250	98	90 - 110	P	07/19/2023	21:03	LB126517
	Cadmium	2480	2500	99	90 - 110	P	07/19/2023	21:03	LB126517
	Chromium	1010	1000	100	90 - 110	P	07/19/2023	21:03	LB126517
	Copper	1250	1250	100	90 - 110	P	07/19/2023	21:03	LB126517
	Lead	5040	5000	101	90 - 110	P	07/19/2023	21:03	LB126517
	Manganese	2460	2500	98	90 - 110	P	07/19/2023	21:03	LB126517
	Nickel	2510	2500	101	90 - 110	P	07/19/2023	21:03	LB126517
	Selenium	5170	5000	103	90 - 110	P	07/19/2023	21:03	LB126517
	Silver	1310	1250	104	90 - 110	P	07/19/2023	21:03	LB126517
	Zinc	2560	2500	102	90 - 110	P	07/19/2023	21:03	LB126517
CCV10	Arsenic	5270	5000	105	90 - 110	P	07/19/2023	21:19	LB126517
	Barium	10400	10000	104	90 - 110	P	07/19/2023	21:19	LB126517
	Beryllium	249	250	100	90 - 110	P	07/19/2023	21:19	LB126517
	Cadmium	2510	2500	100	90 - 110	P	07/19/2023	21:19	LB126517
	Chromium	1020	1000	102	90 - 110	P	07/19/2023	21:19	LB126517
	Copper	1260	1250	101	90 - 110	P	07/19/2023	21:19	LB126517

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Lead	5110	5000	102	90 - 110	P	07/19/2023	21:19	LB126517
	Manganese	2500	2500	100	90 - 110	P	07/19/2023	21:19	LB126517
	Nickel	2550	2500	102	90 - 110	P	07/19/2023	21:19	LB126517
	Selenium	5260	5000	105	90 - 110	P	07/19/2023	21:19	LB126517
	Silver	1320	1250	106	90 - 110	P	07/19/2023	21:19	LB126517
	Zinc	2600	2500	104	90 - 110	P	07/19/2023	21:19	LB126517

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645

Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1080	1000	108	90 - 110	P	07/31/2023	13:25	LB126674
	Barium	530	520	102	90 - 110	P	07/31/2023	13:25	LB126674
	Beryllium	525	510	103	90 - 110	P	07/31/2023	13:25	LB126674
	Cadmium	541	510	106	90 - 110	P	07/31/2023	13:25	LB126674
	Chromium	537	520	103	90 - 110	P	07/31/2023	13:25	LB126674
	Copper	530	510	104	90 - 110	P	07/31/2023	13:25	LB126674
	Lead	1060	1000	106	90 - 110	P	07/31/2023	13:25	LB126674
	Manganese	533	520	102	90 - 110	P	07/31/2023	13:25	LB126674
	Nickel	525	530	99	90 - 110	P	07/31/2023	13:25	LB126674
	Selenium	1090	1000	109	90 - 110	P	07/31/2023	13:25	LB126674
	Silver	259	250	104	90 - 110	P	07/31/2023	13:25	LB126674
	Zinc	1050	1000	105	90 - 110	P	07/31/2023	13:25	LB126674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLCCV01	Arsenic	17.2	20.0	86	80 - 120	P	07/31/2023	13:42	LB126674
	Barium	82.3	100	82	80 - 120	P	07/31/2023	13:42	LB126674
	Beryllium	5.13	6.0	86	80 - 120	P	07/31/2023	13:42	LB126674
	Cadmium	5.33	6.0	89	80 - 120	P	07/31/2023	13:42	LB126674
	Chromium	8.53	10.0	85	80 - 120	P	07/31/2023	13:42	LB126674
	Copper	16.1	20.0	80	80 - 120	P	07/31/2023	13:42	LB126674
	Lead	10.7	12.0	90	80 - 120	P	07/31/2023	13:42	LB126674
	Manganese	17.5	20.0	88	80 - 120	P	07/31/2023	13:42	LB126674
	Nickel	32.2	40.0	80	80 - 120	P	07/31/2023	13:42	LB126674
	Selenium	16.8	20.0	84	80 - 120	P	07/31/2023	13:42	LB126674
	Silver	8.41	10.0	84	80 - 120	P	07/31/2023	13:42	LB126674
	Zinc	34.9	40.0	87	80 - 120	P	07/31/2023	13:42	LB126674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5150	5000	103	90 - 110	P	07/31/2023	14:03	LB126674
	Barium	10300	10000	103	90 - 110	P	07/31/2023	14:03	LB126674
	Beryllium	246	250	98	90 - 110	P	07/31/2023	14:03	LB126674
	Cadmium	2500	2500	100	90 - 110	P	07/31/2023	14:03	LB126674
	Chromium	1010	1000	101	90 - 110	P	07/31/2023	14:03	LB126674
	Copper	1230	1250	98	90 - 110	P	07/31/2023	14:03	LB126674
	Lead	5050	5000	101	90 - 110	P	07/31/2023	14:03	LB126674
	Manganese	2500	2500	100	90 - 110	P	07/31/2023	14:03	LB126674
	Nickel	2520	2500	101	90 - 110	P	07/31/2023	14:03	LB126674
	Selenium	5170	5000	103	90 - 110	P	07/31/2023	14:03	LB126674
	Silver	1290	1250	103	90 - 110	P	07/31/2023	14:03	LB126674
	Zinc	2580	2500	103	90 - 110	P	07/31/2023	14:03	LB126674
CCV02	Arsenic	5100	5000	102	90 - 110	P	07/31/2023	16:21	LB126674
	Barium	10100	10000	102	90 - 110	P	07/31/2023	16:21	LB126674
	Beryllium	232	250	93	90 - 110	P	07/31/2023	16:21	LB126674
	Cadmium	2410	2500	96	90 - 110	P	07/31/2023	16:21	LB126674
	Chromium	972	1000	97	90 - 110	P	07/31/2023	16:21	LB126674
	Copper	1210	1250	97	90 - 110	P	07/31/2023	16:21	LB126674
	Lead	4910	5000	98	90 - 110	P	07/31/2023	16:21	LB126674
	Manganese	2420	2500	97	90 - 110	P	07/31/2023	16:21	LB126674
	Nickel	2450	2500	98	90 - 110	P	07/31/2023	16:21	LB126674
	Selenium	5090	5000	102	90 - 110	P	07/31/2023	16:21	LB126674
	Silver	1240	1250	100	90 - 110	P	07/31/2023	16:21	LB126674
	Zinc	2520	2500	101	90 - 110	P	07/31/2023	16:21	LB126674
CCV03	Arsenic	5220	5000	104	90 - 110	P	07/31/2023	17:09	LB126674
	Barium	10500	10000	105	90 - 110	P	07/31/2023	17:09	LB126674
	Beryllium	243	250	97	90 - 110	P	07/31/2023	17:09	LB126674
	Cadmium	2480	2500	99	90 - 110	P	07/31/2023	17:09	LB126674
	Chromium	1000	1000	100	90 - 110	P	07/31/2023	17:09	LB126674
	Copper	1240	1250	100	90 - 110	P	07/31/2023	17:09	LB126674
	Lead	5030	5000	101	90 - 110	P	07/31/2023	17:09	LB126674
	Manganese	2520	2500	101	90 - 110	P	07/31/2023	17:09	LB126674
	Nickel	2500	2500	100	90 - 110	P	07/31/2023	17:09	LB126674
	Selenium	5190	5000	104	90 - 110	P	07/31/2023	17:09	LB126674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645

Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV03	Silver	1280	1250	102	90 - 110	P	07/31/2023	17:09	LB126674
	Zinc	2560	2500	102	90 - 110	P	07/31/2023	17:09	LB126674
CCV04	Arsenic	4980	5000	100	90 - 110	P	07/31/2023	17:57	LB126674
	Barium	10000	10000	100	90 - 110	P	07/31/2023	17:57	LB126674
	Beryllium	231	250	92	90 - 110	P	07/31/2023	17:57	LB126674
	Cadmium	2360	2500	94	90 - 110	P	07/31/2023	17:57	LB126674
	Chromium	955	1000	96	90 - 110	P	07/31/2023	17:57	LB126674
	Copper	1190	1250	95	90 - 110	P	07/31/2023	17:57	LB126674
	Lead	4780	5000	96	90 - 110	P	07/31/2023	17:57	LB126674
	Manganese	2400	2500	96	90 - 110	P	07/31/2023	17:57	LB126674
	Nickel	2370	2500	95	90 - 110	P	07/31/2023	17:57	LB126674
	Selenium	4980	5000	100	90 - 110	P	07/31/2023	17:57	LB126674
CCV05	Silver	1230	1250	98	90 - 110	P	07/31/2023	17:57	LB126674
	Zinc	2450	2500	98	90 - 110	P	07/31/2023	17:57	LB126674
	Arsenic	5070	5000	101	90 - 110	P	07/31/2023	18:44	LB126674
	Barium	10100	10000	101	90 - 110	P	07/31/2023	18:44	LB126674
	Beryllium	239	250	96	90 - 110	P	07/31/2023	18:44	LB126674
	Cadmium	2410	2500	97	90 - 110	P	07/31/2023	18:44	LB126674
	Chromium	970	1000	97	90 - 110	P	07/31/2023	18:44	LB126674
	Copper	1210	1250	97	90 - 110	P	07/31/2023	18:44	LB126674
	Lead	4890	5000	98	90 - 110	P	07/31/2023	18:44	LB126674
	Manganese	2430	2500	97	90 - 110	P	07/31/2023	18:44	LB126674
CCV06	Nickel	2440	2500	98	90 - 110	P	07/31/2023	18:44	LB126674
	Selenium	5090	5000	102	90 - 110	P	07/31/2023	18:44	LB126674
	Silver	1240	1250	99	90 - 110	P	07/31/2023	18:44	LB126674
	Zinc	2500	2500	100	90 - 110	P	07/31/2023	18:44	LB126674
	Arsenic	4990	5000	100	90 - 110	P	07/31/2023	19:27	LB126674
	Barium	9800	10000	98	90 - 110	P	07/31/2023	19:27	LB126674
	Beryllium	231	250	92	90 - 110	P	07/31/2023	19:27	LB126674
	Cadmium	2370	2500	95	90 - 110	P	07/31/2023	19:27	LB126674
	Chromium	957	1000	96	90 - 110	P	07/31/2023	19:27	LB126674
	Copper	1160	1250	93	90 - 110	P	07/31/2023	19:27	LB126674
	Lead	4810	5000	96	90 - 110	P	07/31/2023	19:27	LB126674
	Manganese	2370	2500	95	90 - 110	P	07/31/2023	19:27	LB126674

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV06	Nickel	2400	2500	96	90 - 110	P	07/31/2023	19:27	LB126674
	Selenium	5010	5000	100	90 - 110	P	07/31/2023	19:27	LB126674
	Silver	1250	1250	100	90 - 110	P	07/31/2023	19:27	LB126674
	Zinc	2470	2500	99	90 - 110	P	07/31/2023	19:27	LB126674
