



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	98	40 - 160	CV	07/19/2023	09:59	LB126505
CRA	Mercury	0.17	0.2	87	40 - 160	CV	07/19/2023	11:58	LB126507
CRI01	Arsenic	15.7	20.0	78	40 - 160	P	07/19/2023	13:17	LB126517
	Barium	93.7	100	94	40 - 160	P	07/19/2023	13:17	LB126517
	Beryllium	5.81	6.0	97	40 - 160	P	07/19/2023	13:17	LB126517
	Cadmium	6.03	6.0	100	40 - 160	P	07/19/2023	13:17	LB126517
	Chromium	8.98	10.0	90	40 - 160	P	07/19/2023	13:17	LB126517
	Copper	20.5	20.0	103	40 - 160	P	07/19/2023	13:17	LB126517
	Lead	13.6	12.0	113	40 - 160	P	07/19/2023	13:17	LB126517
	Manganese	20.4	20.0	102	40 - 160	P	07/19/2023	13:17	LB126517
	Nickel	36.2	40.0	91	40 - 160	P	07/19/2023	13:17	LB126517
	Selenium	20.2	20.0	101	40 - 160	P	07/19/2023	13:17	LB126517
	Silver	10.1	10.0	102	40 - 160	P	07/19/2023	13:17	LB126517
	Zinc	39.4	40.0	99	40 - 160	P	07/19/2023	13:17	LB126517
CRI01	Arsenic	22.1	20.0	110	40 - 160	P	07/31/2023	13:51	LB126674
	Barium	96.4	100	96	40 - 160	P	07/31/2023	13:51	LB126674
	Beryllium	5.86	6.0	98	40 - 160	P	07/31/2023	13:51	LB126674
	Cadmium	6.25	6.0	104	40 - 160	P	07/31/2023	13:51	LB126674
	Chromium	9.68	10.0	97	40 - 160	P	07/31/2023	13:51	LB126674
	Copper	20.1	20.0	101	40 - 160	P	07/31/2023	13:51	LB126674
	Lead	11.9	12.0	100	40 - 160	P	07/31/2023	13:51	LB126674
	Manganese	20.3	20.0	102	40 - 160	P	07/31/2023	13:51	LB126674
	Nickel	37.9	40.0	95	40 - 160	P	07/31/2023	13:51	LB126674
	Selenium	24.4	20.0	122	40 - 160	P	07/31/2023	13:51	LB126674
	Silver	10.4	10.0	104	40 - 160	P	07/31/2023	13:51	LB126674
	Zinc	40.8	40.0	102	40 - 160	P	07/31/2023	13:51	LB126674