

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	-3.33			-20	20	07/19/2023	13:21	LB126517
	Barium	5.90	6.0	98	-94	106	07/19/2023	13:21	LB126517
	Beryllium	1.04			-6	6	07/19/2023	13:21	LB126517
	Cadmium	1.04	1.0	104	-5	7	07/19/2023	13:21	LB126517
	Chromium	48.5	52.0	93	42	62	07/19/2023	13:21	LB126517
	Copper	-2.10	2.0	105	-18	22	07/19/2023	13:21	LB126517
	Lead	-2.93			-12	12	07/19/2023	13:21	LB126517
	Manganese	15.9	7.0	227	-13	27	07/19/2023	13:21	LB126517
	Nickel	0.98	2.0	49	-38	42	07/19/2023	13:21	LB126517
	Selenium	-2.32			-20	20	07/19/2023	13:21	LB126517
	Silver	-2.78			-10	10	07/19/2023	13:21	LB126517
	Zinc	4.30			-40	40	07/19/2023	13:21	LB126517
ICSAB01	Arsenic	102	100	102	85	120	07/19/2023	13:25	LB126517
	Barium	537	540	99	440	640	07/19/2023	13:25	LB126517
	Beryllium	509	500	102	430	580	07/19/2023	13:25	LB126517
	Cadmium	1060	970	109	820	1100	07/19/2023	13:25	LB126517
	Chromium	562	540	104	460	620	07/19/2023	13:25	LB126517
	Copper	505	510	99	430	590	07/19/2023	13:25	LB126517
	Lead	56.7	49.0	116	37	61	07/19/2023	13:25	LB126517
	Manganese	525	510	103	430	590	07/19/2023	13:25	LB126517
	Nickel	1030	950	108	810	1100	07/19/2023	13:25	LB126517
	Selenium	44.6	46.0	97	26	66	07/19/2023	13:25	LB126517
	Silver	177	200	88	170	230	07/19/2023	13:25	LB126517
	Zinc	1060	950	112	810	1100	07/19/2023	13:25	LB126517
ICSA01	Arsenic	-19.9			-20	20	07/31/2023	13:55	LB126674
	Barium	5.67	6.0	94	-94	106	07/31/2023	13:55	LB126674
	Beryllium	0.96			-6	6	07/31/2023	13:55	LB126674
	Cadmium	1.09	1.0	109	-5	7	07/31/2023	13:55	LB126674
	Chromium	49.1	52.0	94	42	62	07/31/2023	13:55	LB126674
	Copper	6.72	2.0	336	-18	22	07/31/2023	13:55	LB126674
	Lead	-1.35			-12	12	07/31/2023	13:55	LB126674
	Manganese	12.2	7.0	174	-13	27	07/31/2023	13:55	LB126674
	Nickel	0.53	2.0	26	-38	42	07/31/2023	13:55	LB126674
	Selenium	19.5			-20	20	07/31/2023	13:55	LB126674
	Silver	-2.56			-10	10	07/31/2023	13:55	LB126674
	Zinc	2.74			-40	40	07/31/2023	13:55	LB126674
ICSAB01	Arsenic	90.1	100	90	85	120	07/31/2023	13:59	LB126674
	Barium	524	540	97	440	640	07/31/2023	13:59	LB126674
	Beryllium	502	500	100	430	580	07/31/2023	13:59	LB126674
	Cadmium	1050	970	108	820	1100	07/31/2023	13:59	LB126674
	Chromium	560	540	104	460	620	07/31/2023	13:59	LB126674
	Copper	504	510	99	430	590	07/31/2023	13:59	LB126674

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
ICS Source: EPA **Instrument ID:** P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Lead	53.6	49.0	109	37	61	07/31/2023	13:59	LB126674
	Manganese	511	510	100	430	590	07/31/2023	13:59	LB126674
	Nickel	1020	950	107	810	1100	07/31/2023	13:59	LB126674
	Selenium	60.4	46.0	131	26	66	07/31/2023	13:59	LB126674
	Silver	218	200	109	170	230	07/31/2023	13:59	LB126674
	Zinc	1060	950	112	810	1100	07/31/2023	13:59	LB126674