

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LaBella Associates P.C. **Level:** LOW **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Matrix: Water **Sample ID:** O3637-01 **Client ID:** A508DUP
Percent Solids for Sample: NA **Duplicate ID** O3637-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	13.2		15.4		15		P
Barium	ug/L	20	383		381		1		P
Beryllium	ug/L	20	0.64	J	0.62	J	4		P
Cadmium	ug/L	20	1.40	J	1.40	J	0		P
Chromium	ug/L	20	59.9		60.7		1		P
Copper	ug/L	20	58.4		59.1		1		P
Lead	ug/L	20	240		244		2		P
Manganese	ug/L	20	1180		1180		0		P
Mercury	ug/L	20	2.93		2.48		17		CV
Nickel	ug/L	20	18.0	J	18.1	J	1		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Zinc	ug/L	20	337		342		1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LaBella Associates P.C. **Level:** LOW **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Matrix: Water **Sample ID:** O3637-01MS **Client ID:** A508MSD
Percent Solids for Sample: NA **Duplicate ID** O3637-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	436		430		1		P
Barium	ug/L	20	430		423		2		P
Beryllium	ug/L	20	91.3		90.9		0		P
Cadmium	ug/L	20	99.7		98.2		2		P
Chromium	ug/L	20	246		242		2		P
Copper	ug/L	20	198		195		2		P
Lead	ug/L	20	706		693		2		P
Manganese	ug/L	20	1090		1070		2		P
Mercury	ug/L	20	6.37		6.40		0		CV
Nickel	ug/L	20	259		255		2		P
Selenium	ug/L	20	941		921		2		P
Silver	ug/L	20	37.4		36.7		2		P
Zinc	ug/L	20	401		390		3		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LaBella Associates P.C. **Level:** LOW **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Matrix: Solid **Sample ID:** O3645-02 **Client ID:** SB-04-(1-5)DUP
Percent Solids for Sample: 91.3 **Duplicate ID** O3645-02DUP **Percent Solids for Spike Sample:** 91.3

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	mg/Kg	20	1.15		0.98		16		P
Barium	mg/Kg	20	27.7		27.2		2		P
Beryllium	mg/Kg	20	0.23	J	0.23	J	2		P
Cadmium	mg/Kg	20	1.81		1.81		0		P
Chromium	mg/Kg	20	4.23		4.28		1		P
Copper	mg/Kg	20	15.6		15.3		2		P
Lead	mg/Kg	20	84.4		84.7		0		P
Manganese	mg/Kg	20	483		472		2		P
Mercury	mg/Kg	20	6.05	D	6.66	D	10		CV
Nickel	mg/Kg	20	9.56		9.59		0		P
Selenium	mg/Kg	20	0.90	U	0.90	U			P
Silver	mg/Kg	20	0.45	U	0.45	U			P
Zinc	mg/Kg	20	314		318		1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LaBella Associates P.C. **Level:** LOW **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645
Matrix: Solid **Sample ID:** O3645-09 **Client ID:** SB-04-(1-5)MSD
Percent Solids for Sample: 91.3 **Duplicate ID** O3645-10 **Percent Solids for Spike Sample:** 91.3

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	mg/Kg	20	35.0		34.0		3		P
Barium	mg/Kg	20	35.4		40.4		13		P
Beryllium	mg/Kg	20	6.97		6.81		2		P
Cadmium	mg/Kg	20	10.6		10.3		3		P
Chromium	mg/Kg	20	17.4		16.9		3		P
Copper	mg/Kg	20	26.9		34.0		23	*	P
Lead	mg/Kg	20	133		133		0		P
Manganese	mg/Kg	20	490		496		1		P
Mercury	mg/Kg	20	6.41	D	6.17	D	4		CV
Nickel	mg/Kg	20	31.2		29.2		7		P
Selenium	mg/Kg	20	73.5		72.2		2		P
Silver	mg/Kg	20	2.83		2.80		1		P
Zinc	mg/Kg	20	330		299		10		P