

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LaBella Associates P.C.

SDG No.: O3645

Contract: LABE01

Lab Code: CHEM

Case No.: O3645

SAS No.: O3645

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB154230BS							
Arsenic	ug/L	400	414		104	80 - 120	P
Barium	ug/L	100	100		100	80 - 120	P
Beryllium	ug/L	100	94.8		95	80 - 120	P
Cadmium	ug/L	100	98.2		98	80 - 120	P
Chromium	ug/L	200	197		98	80 - 120	P
Copper	ug/L	150	154		103	80 - 120	P
Lead	ug/L	500	490		98	80 - 120	P
Manganese	ug/L	100	100		100	80 - 120	P
Nickel	ug/L	250	240		96	80 - 120	P
Selenium	ug/L	1000	1010		101	80 - 120	P
Silver	ug/L	37.5	37.4		100	80 - 120	P
Zinc	ug/L	100	102		102	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LaBella Associates P.C. **SDG No.:** O3645
Contract: LABE01 **Lab Code:** CHEM **Case No.:** O3645 **SAS No.:** O3645

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB154232BS							
Arsenic	mg/Kg	36.7	38.4		105	80 - 120	P
Barium	mg/Kg	9.2	9.13		99	80 - 120	P
Beryllium	mg/Kg	9.2	8.60		94	80 - 120	P
Cadmium	mg/Kg	9.2	8.88		96	80 - 120	P
Chromium	mg/Kg	18.3	18.0		98	80 - 120	P
Copper	mg/Kg	13.8	14.2		103	80 - 120	P
Lead	mg/Kg	45.9	44.4		97	80 - 120	P
Manganese	mg/Kg	9.2	8.94		97	80 - 120	P
Nickel	mg/Kg	22.9	21.9		96	80 - 120	P
Selenium	mg/Kg	91.7	93.3		102	80 - 120	P
Silver	mg/Kg	3.4	3.41		100	80 - 120	P
Zinc	mg/Kg	9.2	9.36		102	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LaBella Associates P.C.

SDG No.: O3645

Contract: LABE01

Lab Code: CHEM

Case No.: O3645

SAS No.: O3645

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB154278BS Mercury	mg/Kg	0.25	0.23		92	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LaBella Associates P.C.

SDG No.: O3645

Contract: LABE01

Lab Code: CHEM

Case No.: O3645

SAS No.: O3645

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB154279BS Mercury	ug/L	4.0	3.58		90	80 - 120	CV
