



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: LABE01

Lab Code: CHEM Case No.: O3645 SAS No.: O3645 SDG NO.: O3645

Continuing Calib Date: 07/18/2023 Initial Calibration Date(s): 07/01/2023 07/01/2023

Continuing Calib Time: 11:01 Initial Calibration Time(s): 03:04 09:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.50	4.50	4.40	4.60	0.00
Aroclor-1242-2 (2)	4.52	4.52	4.42	4.62	0.00
Aroclor-1242-3 (3)	4.58	4.58	4.48	4.68	0.00
Aroclor-1242-4 (4)	4.67	4.67	4.57	4.77	0.00
Aroclor-1242-5 (5)	5.39	5.39	5.29	5.49	0.00
Tetrachloro-m-xylene	3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl	8.58	8.58	8.48	8.68	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: LABE01

Lab Code: CHEM Case No.: O3645 SAS No.: O3645 SDG NO.: O3645

Continuing Calib Date: 07/18/2023 Initial Calibration Date(s): 07/01/2023 07/01/2023

Continuing Calib Time: 11:01 Initial Calibration Time(s): 03:04 09:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	3.76	3.76	3.66	3.86	0.00
Aroclor-1242-2 (2)	3.77	3.78	3.68	3.88	0.01
Aroclor-1242-3 (3)	3.93	3.94	3.84	4.04	0.01
Aroclor-1242-4 (4)	4.02	4.02	3.92	4.12	0.00
Aroclor-1242-5 (5)	4.51	4.52	4.42	4.62	0.01
Tetrachloro-m-xylene	2.76	2.76	2.66	2.86	0.00
Decachlorobiphenyl	7.53	7.53	7.43	7.63	0.00