



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: LABE01

Lab Code: CHEM Case No.: O3645 SAS No.: O3645 SDG NO.: O3645

Continuing Calib Date: 07/18/2023 Initial Calibration Date(s): 07/01/2023 07/01/2023

Continuing Calib Time: 21:53 Initial Calibration Time(s): 03:04 09:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.50	4.50	4.40	4.60	0.00
Aroclor-1016-2	(2)	4.52	4.52	4.42	4.62	0.00
Aroclor-1016-3	(3)	4.58	4.58	4.48	4.68	0.00
Aroclor-1016-4	(4)	4.67	4.67	4.57	4.77	0.00
Aroclor-1016-5	(5)	4.96	4.96	4.86	5.06	0.00
Aroclor-1260-1	(1)	6.06	6.06	5.96	6.16	0.00
Aroclor-1260-2	(2)	6.32	6.32	6.22	6.42	0.00
Aroclor-1260-3	(3)	6.68	6.68	6.58	6.78	0.00
Aroclor-1260-4	(4)	6.89	6.89	6.79	6.99	0.00
Aroclor-1260-5	(5)	7.20	7.20	7.10	7.30	0.00
Tetrachloro-m-xylene		3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl		8.58	8.58	8.48	8.68	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: LABE01

Lab Code: CHEM Case No.: O3645 SAS No.: O3645 SDG NO.: O3645

Continuing Calib Date: 07/18/2023 Initial Calibration Date(s): 07/01/2023 07/01/2023

Continuing Calib Time: 21:53 Initial Calibration Time(s): 03:04 09:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	3.76	3.76	3.66	3.86	0.00
Aroclor-1016-2	(2)	3.77	3.78	3.68	3.88	0.01
Aroclor-1016-3	(3)	3.93	3.94	3.84	4.04	0.01
Aroclor-1016-4	(4)	3.98	3.99	3.89	4.09	0.01
Aroclor-1016-5	(5)	4.18	4.18	4.08	4.28	0.00
Aroclor-1260-1	(1)	5.17	5.17	5.07	5.27	0.00
Aroclor-1260-2	(2)	5.36	5.36	5.26	5.46	0.00
Aroclor-1260-3	(3)	5.50	5.50	5.40	5.60	0.00
Aroclor-1260-4	(4)	5.96	5.97	5.87	6.07	0.01
Aroclor-1260-5	(5)	6.21	6.21	6.11	6.31	0.00
Tetrachloro-m-xylene		2.76	2.76	2.66	2.86	0.00
Decachlorobiphenyl		7.52	7.53	7.43	7.63	0.01