

Report of Analysis

Client:	LaBella Associates P.C.	Date Collected:	07/19/23
Project:	Mackenna Parcels	Date Received:	07/19/23
Client Sample ID:	PIBLK-PQ062377.D	SDG No.:	O3645
Lab Sample ID:	I.BLK-PQ062377.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB Group1
Extraction Type:		Injection Volume :	
GPC Factor :	1.0	PH :	
Prep Method :	5030		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PQ062377.D	1		07/19/23	PQ071823

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.50	U	0.15	0.50	ug/L
11104-28-2	Aroclor-1221	0.50	U	0.22	0.50	ug/L
11141-16-5	Aroclor-1232	0.50	U	0.18	0.50	ug/L
53469-21-9	Aroclor-1242	0.50	U	0.18	0.50	ug/L
12672-29-6	Aroclor-1248	0.50	U	0.15	0.50	ug/L
11097-69-1	Aroclor-1254	0.50	U	0.15	0.50	ug/L
11096-82-5	Aroclor-1260	0.50	U	0.16	0.50	ug/L
37324-23-5	Aroclor-1262	0.50	U	0.16	0.50	ug/L
11100-14-4	Aroclor-1268	0.50	U	0.13	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.9		60 - 140	84%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.1		60 - 140	96%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit