

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072423\
 Data File : BF134435.D
 Acq On : 24 Jul 2023 18:16
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF072423

Quant Time: Jul 25 01:47:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 25 01:46:04 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	111	0.00
2	1,4-Dioxane	40.000	39.007	2.5	110	0.00
3	Pyridine	40.000	43.386	-8.5	114	0.00
4	n-Nitrosodimethylamine	40.000	42.927	-7.3	118	0.00
5 S	2-Fluorophenol	80.000	82.473	-3.1	110	0.00
6	Aniline	40.000	40.828	-2.1	111	0.00
7 S	Phenol-d6	80.000	80.455	-0.6	112	0.00
8	2-Chlorophenol	40.000	40.367	-0.9	110	0.00
9	Benzaldehyde	40.000	38.639	3.4	114	0.00
10 C	Phenol	40.000	40.129	-0.3	110	0.00
11	bis(2-Chloroethyl)ether	40.000	40.002	-0.0	112	0.00
12	1,3-Dichlorobenzene	40.000	39.945	0.1	107	0.00
13 C	1,4-Dichlorobenzene	40.000	40.566	-1.4	111	0.00
14	1,2-Dichlorobenzene	40.000	40.116	-0.3	112	0.00
15	Benzyl Alcohol	40.000	40.137	-0.3	105	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.906	-2.3	114	0.00
17	2-Methylphenol	40.000	39.121	2.2	106	0.00
18	Hexachloroethane	40.000	39.159	2.1	110	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.681	3.3	107	0.00
20	3+4-Methylphenols	40.000	39.291	1.8	108	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	110	0.00
22	Acetophenone	40.000	40.320	-0.8	110	0.00
23 S	Nitrobenzene-d5	80.000	83.445	-4.3	112	0.00
24	Nitrobenzene	40.000	40.867	-2.2	113	0.00
25	Isophorone	40.000	40.103	-0.3	109	0.00
26 C	2-Nitrophenol	40.000	40.952	-2.4	109	0.00
27	2,4-Dimethylphenol	40.000	41.429	-3.6	113	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.472	-1.2	111	0.00
29 C	2,4-Dichlorophenol	40.000	40.572	-1.4	109	0.00
30	1,2,4-Trichlorobenzene	40.000	40.054	-0.1	108	0.00
31	Naphthalene	40.000	40.147	-0.4	109	0.00
32	Benzoic acid	40.000	41.715	-4.3	108	0.00
33	4-Chloroaniline	40.000	40.558	-1.4	109	0.00
34 C	Hexachlorobutadiene	40.000	40.192	-0.5	108	0.00
35	Caprolactam	40.000	40.785	-2.0	108	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.350	-0.9	109	0.00
37	2-Methylnaphthalene	40.000	40.005	-0.0	108	0.00
38	1-Methylnaphthalene	40.000	40.115	-0.3	109	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.968	-2.4	108	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.716	-4.3	109	0.00
42 S	2,4,6-Tribromophenol	80.000	82.241	-2.8	97	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.531	-1.3	96	0.00
44	2,4,5-Trichlorophenol	40.000	40.214	-0.5	94	0.00
45 S	2-Fluorobiphenyl	80.000	80.759	-0.9	96	0.00
46	1,1'-Biphenyl	40.000	40.371	-0.9	96	0.00
47	2-Chloronaphthalene	40.000	40.366	-0.9	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072423\
 Data File : BF134435.D
 Acq On : 24 Jul 2023 18:16
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF072423

Quant Time: Jul 25 01:47:04 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072423.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 25 01:46:04 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.063	-5.2	97	0.00
49	Acenaphthylene	40.000	40.216	-0.5	95	0.00
50	Dimethylphthalate	40.000	40.055	-0.1	94	0.00
51	2,6-Dinitrotoluene	40.000	40.030	-0.1	93	0.00
52 C	Acenaphthene	40.000	41.070	-2.7	96	0.00
53	3-Nitroaniline	40.000	41.598	-4.0	97	0.00
54 P	2,4-Dinitrophenol	40.000	37.380	6.5	93	0.00
55	Dibenzofuran	40.000	40.272	-0.7	95	0.00
56 P	4-Nitrophenol	40.000	39.250	1.9	92	0.00
57	2,4-Dinitrotoluene	40.000	41.664	-4.2	95	0.00
58	Fluorene	40.000	40.390	-1.0	95	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.536	-1.3	95	0.00
60	Diethylphthalate	40.000	40.064	-0.2	95	0.00
61	4-Chlorophenyl-phenylether	40.000	40.476	-1.2	95	0.00
62	4-Nitroaniline	40.000	41.397	-3.5	95	0.00
63	Azobenzene	40.000	40.522	-1.3	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.807	-2.0	93	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.214	-0.5	94	0.00
67	4-Bromophenyl-phenylether	40.000	39.890	0.3	94	0.00
68	Hexachlorobenzene	40.000	40.380	-1.0	96	0.00
69	Atrazine	40.000	41.716	-4.3	96	0.00
70 C	Pentachlorophenol	40.000	40.604	-1.5	94	0.00
71	Phenanthrene	40.000	40.194	-0.5	94	0.00
72	Anthracene	40.000	40.715	-1.8	95	0.00
73	Carbazole	40.000	40.918	-2.3	97	0.00
74	Di-n-butylphthalate	40.000	39.947	0.1	93	0.00
75 C	Fluoranthene	40.000	40.955	-2.4	97	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	95	0.00
77	Benzydine	40.000	38.592	3.5	89	0.00
78	Pyrene	40.000	40.700	-1.8	96	0.00
79 S	Terphenyl-d14	80.000	80.774	-1.0	96	0.00
80	Butylbenzylphthalate	40.000	40.971	-2.4	95	0.00
81	Benzo(a)anthracene	40.000	39.849	0.4	96	0.00
82	3,3'-Dichlorobenzidine	40.000	38.241	4.4	91	0.00
83	Chrysene	40.000	40.820	-2.1	96	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.980	-2.4	96	0.00
85 c	Di-n-octyl phthalate	40.000	40.155	-0.4	94	0.00
86 I	Perylene-d12	20.000	20.000	0.0	97	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.890	-4.7	96	-0.01
88	Benzo(b)fluoranthene	40.000	38.814	3.0	96	0.00
89	Benzo(k)fluoranthene	40.000	41.555	-3.9	96	0.00
90 C	Benzo(a)pyrene	40.000	40.864	-2.2	97	0.00
91	Dibenzo(a,h)anthracene	40.000	42.108	-5.3	97	0.00
92	Benzo(g,h,i)perylene	40.000	42.833	-7.1	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF072423\
Data File : BF134435.D
Acq On : 24 Jul 2023 18:16
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF072423

Quant Time: Jul 25 01:47:04 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF072423.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 25 01:46:04 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0