



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080123\
Data File : BF134535.D
Acq On : 01 Aug 2023 07:20
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Aug 02 02:20:34 2023






AutoFind: Scans 1613, 1614, 1615; Background Corrected with Scan 1607

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.8	85351	PASS
68	69	0.00	2	1.8	1385	PASS
69	198	0.00	100	33.0	76597	PASS
70	69	0.00	2	0.7	501	PASS
127	198	10	80	44.7	103757	PASS
197	198	0.00	2	0.8	1744	PASS
198	198	100	100	100.0	232184	PASS
199	198	5	9	6.8	15885	PASS
275	198	10	60	27.2	63099	PASS
365	198	1	100	3.9	9066	PASS
441	198	0.01	100	17.1	39813	PASS
442	442	50	100	100.0	260459	PASS
443	442	15	24	18.8	49024	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080123\
Data File : BF134535.D
Acq On : 01 Aug 2023 07:20
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Aug 02 02:26:34 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 02 02:20:34 2023
Response via : Initial Calibration



TIC: BF134535.D\data.ms

(70) Pentachlorophenol (C)

11.110min (-0.006) 31442.83 ng




response 218296

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	66.10	64.63
264.00	59.70	63.59
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF080123\
Data File : BF134535.D
Acq On : 01 Aug 2023 07:20
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Aug 02 02:26:34 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF080123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 02 02:20:34 2023
Response via : Initial Calibration



TIC: BF134535.D\data.ms

(77) Benzidine

12.657min (+ 0.000) 0.00 ng

response 1518267

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.00	14.40
183.00	10.80	11.72
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/1/2023	BNA_F	<u>BF134535.D</u>
Compound Name	Response	Retention Time
DDT	422953	13.48
DDD	3615	13.21
DDE		12.845
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3615	426568	0.85

Instrument :
BNA_F
ClientSampleId :
DFTPP