

### NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

## GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NU	MBER:	BF080123			
SequenceID :	BF080123		NA	NO	YES
1. Chromatograms Labeled/C	Compounds Identif	ied. (Field samples and Method Blanks)			✓
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria   (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)					✓
3. GC/MS Tuning Frequency series	- Performed every	24 hours for 600 series and 12 hours for 8000			✓
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series				✓	
5. GC/MS Calibration Met:					✓
a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.				_	
U		l Calibration method (Method 8270-BF080123.M) for 2-Nitrop Dinitrotoluene ,4,6-Dinitro-2-methylphenol, this compound is j			
b. Continuous Calibration( If not met, list those compour		ia veries which fall outside the acceptable range.			✓
6. Blank Contamination - If y	ves, list compound	s and concentrations in each blank:		_✓	

- a. B/N Fraction
- d. Acid Fraction

7. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			<u>✓</u>
a. B/N Fraction			
d. Acid Fraction			
8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	<u>✓</u>		
a. B/N Fraction			
d. Acid Fraction			
9. Internal Standard Area/Retention Time Shift Meet Criteria Comments:		<u>✓</u>	
One Internal Standard is bias low in O3784-01, sample is very bad matrix and having parabolic type chromato, further analysis for this sample, this run will be used as final run in hardcopy.	gram, no		
10. Extraction Holding Time Met If not met, list number of days exceeded for each sample:	<u>✓</u>		
11. Analysis Holding Time Met If not met, list number of days exceeded for each sample:	<u> </u>		

### ADDITIONAL COMMENTS:

One Internal Standard is bias low in O3784-01, sample is very bad matrix and having parabolic type chromatogram, no further analysis for this sample, this run will be used as final run in hardcopy.



### NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

# GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER:

SequenceID :

BF080123

BF080123

NA NO

YES