



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: O3645 SAS No.: O3645 SDG NO.: O3645

EPA Sample No.: SSTDCCC040 Date Analyzed: 07/19/2023

Lab File ID: BF134355.D Time Analyzed: 13:01

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	72147	6.84	261643	8.12	134896	9.88
UPPER LIMIT	144294	7.34	523286	8.622	269792	10.381
LOWER LIMIT	36073.5	6.34	130822	7.622	67448	9.381
EPA SAMPLE NO.						
01 PB154258BL	76908	6.84	293261	8.12	157065	9.88
02 PB154258BS	73456	6.84	287399	8.12	147548	9.88

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.