



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: 03645 SAS No.: 03645 SDG NO.: 03645

EPA Sample No.: SSTDCCC040 Date Analyzed: 07/28/2023

Lab File ID: BF134518.D Time Analyzed: 16:29

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	87669	6.828	337479	8.11	185549	9.87
UPPER LIMIT	175338	7.328	674958	8.61	371098	10.369
LOWER LIMIT	43834.5	6.328	168740	7.61	92774.5	9.369
EPA SAMPLE NO.						
01 SB-02-(3-5)	80491	6.83	305901	8.11	133240	9.86
02 SB-07-(1-3)	99614	6.83	368068	8.11	164170	9.86
03 RINSATE-BLANK	116133	6.83	261936	8.11	136997	9.86

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.