



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH
Lab Code: CHEM Case No.: 03645 SAS No.: 03645 SDG NO.: 03645
EPA Sample No.: SSTDCCC040 Date Analyzed: 08/01/2023
Lab File ID: BF134547.D Time Analyzed: 14:27
Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	220485	6.792	863869	8.08	433432	9.83
UPPER LIMIT	440970	7.292	1727740	8.575	866864	10.328
LOWER LIMIT	110243	6.292	431935	7.575	216716	9.328
EPA SAMPLE NO.						
01 SB-04-(1-5)	205461	6.79	825053	8.07	411202	9.82
02 SB-08-(0.5-2.0)	186884	6.79	736666	8.07	365569	9.82
03 SB-10-(0.5-2.0)	186419	6.79	688439	8.07	288810	9.83
04 SB-10-(0.5-2.0)DL	176047	6.79	669108	8.07	288095	9.82

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.