



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: O3645 SAS No.: O3645 SDG NO.: O3645

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/02/2023

Lab File ID: BG058569.D Time Analyzed: 08:34

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	39596	7.814	168505	10.62	117751	14.47
UPPER LIMIT	79192	8.314	337010	11.116	235502	14.965
LOWER LIMIT	19798	7.314	84252.5	10.116	58875.5	13.965
EPA SAMPLE NO.						
01 DUP	54082	7.81	234812	10.61	164688	14.46
02 SB-09-(2.0-4.0)	58172	7.81	254632	10.62	173135	14.46

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.