



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: O3645 SAS No.: O3645 SDG NO.: O3645

EPA Sample No.: SSTDCCC040 Date Analyzed: 07/28/2023

Lab File ID: BM041169.D Time Analyzed: 17:11

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	155350	7.804	581542	10.62	321278	14.47
UPPER LIMIT	310700	8.304	1163080	11.116	642556	14.969
LOWER LIMIT	77675	7.304	290771	10.116	160639	13.969
EPA SAMPLE NO.						
01 SB-04-(1-5)MS	183644	7.80	715690	10.61	398072	14.47
02 SB-04-(1-5)MSD	193205	7.80	741556	10.61	413329	14.47

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.