



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH
Lab Code: CHEM Case No.: 03645 SAS No.: 03645 SDG NO.: 03645
EPA Sample No.: SSTDCCC040 Date Analyzed: 07/28/2023
Lab File ID: BP016411.D Time Analyzed: 08:03
Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	506538	8.093	2090640	10.92	1248290	14.73
UPPER LIMIT	1013080	8.593	4181280	11.422	2496580	15.234
LOWER LIMIT	253269	7.593	1045320	10.422	624145	14.234
EPA SAMPLE NO.						
01 PB154218BL	431547	8.09	1746470	10.92	1045570	14.73
02 PB154218BS	349249	8.09	1482470	10.92	905687	14.73
03 PB154218BSD	341099	8.09	1443210	10.92	877872	14.73

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.