



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: 03645 SAS No.: 03645 SDG NO.: 03645

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/02/2023

Lab File ID: BG058569.D Time Analyzed: 08:34

Instrument ID: BNA_G GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	312836	17.209	298102	21.457	377370	24.529
UPPER LIMIT	625672	17.709	596204	21.957	754740	25.029
LOWER LIMIT	156418	16.709	149051	20.957	188685	24.029
EPA SAMPLE NO.						
01 DUP	370089	17.21	322252	21.45	402408	24.52
02 SB-09- (2.0-4.0)	388943	17.21	335998	21.45	415042	24.52

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.