



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: 03645 SAS No.: 03645 SDG NO.: 03645

EPA Sample No.: SSTDCCC040 Date Analyzed: 07/28/2023

Lab File ID: BP016411.D Time Analyzed: 08:03

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	2658100	17.498	2292480	21.592	2240140	24.186
	UPPER LIMIT	5316200	17.998	4584960	22.092	4480280	24.686
	LOWER LIMIT	1329050	16.998	1146240	21.092	1120070	23.686
	EPA SAMPLE NO.						
01	PB154218BL	2341400	17.49	2050870	21.59	2195260	24.18
02	PB154218BS	1936050	17.50	1800690	21.59	1921990	24.19
03	PB154218BSD	1902270	17.50	1781260	21.59	1923430	24.19

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.