



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: O3645

Client: LaBella Associates P.C.

Analytical Method: SW8260D

Datafile : VD076768.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VD0718SBS01	Dichlorodifluoromethane	20	23.8	ug/Kg	119			64	136	
	Chloromethane	20	19.2	ug/Kg	96			70	130	
	Vinyl chloride	20	20.0	ug/Kg	100			72	129	
	Bromomethane	20	20.5	ug/Kg	103			58	141	
	Chloroethane	20	20.2	ug/Kg	101			69	130	
	Trichlorofluoromethane	20	22.6	ug/Kg	113			69	134	
	1,1,2-Trichlorotrifluoroethane	20	22.6	ug/Kg	113			81	123	
	1,1-Dichloroethene	20	20.8	ug/Kg	104			79	121	
	Acetone	100	92.4	ug/Kg	92			60	131	
	Carbon disulfide	20	19.6	ug/Kg	98			45	154	
	Methyl tert-butyl Ether	20	18.5	ug/Kg	93			77	129	
	Methyl Acetate	20	20.4	ug/Kg	102			69	149	
	Methylene Chloride	20	21.1	ug/Kg	106			39	175	
	trans-1,2-Dichloroethene	20	19.8	ug/Kg	99			80	123	
	1,1-Dichloroethane	20	20.4	ug/Kg	102			82	123	
	Cyclohexane	20	20.8	ug/Kg	104			76	122	
	2-Butanone	100	98.9	ug/Kg	99			69	131	
	Carbon Tetrachloride	20	21.9	ug/Kg	110			76	129	
	cis-1,2-Dichloroethene	20	19.7	ug/Kg	99			82	123	
	Bromochloromethane	20	22.2	ug/Kg	111			62	134	
	Chloroform	20	20.1	ug/Kg	101			82	125	
	1,1,1-Trichloroethane	20	21.3	ug/Kg	106			80	126	
	Methylcyclohexane	20	22.0	ug/Kg	110			77	123	
	Benzene	20	20.3	ug/Kg	102			84	121	
	1,2-Dichloroethane	20	20.8	ug/Kg	104			81	126	
	Trichloroethene	20	20.0	ug/Kg	100			83	122	
	1,2-Dichloropropane	20	20.6	ug/Kg	103			83	122	
	Bromodichloromethane	20	20.3	ug/Kg	102			82	123	
	4-Methyl-2-Pentanone	100	94.2	ug/Kg	94			70	135	
	Toluene	20	20.0	ug/Kg	100			83	122	
	t-1,3-Dichloropropene	20	19.0	ug/Kg	95			78	124	
	cis-1,3-Dichloropropene	20	19.3	ug/Kg	97			81	122	
	1,1,2-Trichloroethane	20	20.2	ug/Kg	101			82	125	
	2-Hexanone	100	93.8	ug/Kg	94			66	138	
	Dibromochloromethane	20	20.1	ug/Kg	101			79	125	
	1,2-Dibromoethane	20	19.6	ug/Kg	98			80	125	
	Tetrachloroethene	20	23.8	ug/Kg	119			83	125	
	Chlorobenzene	20	20.9	ug/Kg	104			84	122	
	Ethyl Benzene	20	21.1	ug/Kg	106			82	124	
	m/p-Xylenes	40	41.7	ug/Kg	104			83	124	
	o-Xylene	20	20.4	ug/Kg	102			83	123	
	Styrene	20	20.9	ug/Kg	104			82	124	
	Bromoform	20	20.9	ug/Kg	104			75	127	
	Isopropylbenzene	20	20.7	ug/Kg	104			82	124	
	1,1,2,2-Tetrachloroethane	20	20.4	ug/Kg	102			77	127	
	N-propylbenzene	20	21.1	ug/Kg	106			81	123	
	1,3,5-Trimethylbenzene	20	20.8	ug/Kg	104			82	124	
	tert-Butylbenzene	20	21.0	ug/Kg	105			81	125	
	1,2,4-Trimethylbenzene	20	20.2	ug/Kg	101			81	125	



Laboratory Control Sample/Laboratory Control Sample Duplicate Summary
SW-846

SDG No.: O3645

Client: LaBella Associates P.C.

Analytical Method: SW8260D

Datafile : VD076768.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	
									High	RPD
VD0718SBS01	Sec-butylbenzene	20	21.7	ug/Kg	109			80	124	
	p-Isopropyltoluene	20	21.2	ug/Kg	106			81	125	
	1,3-Dichlorobenzene	20	19.9	ug/Kg	100			83	122	
	1,4-Dichlorobenzene	20	20.5	ug/Kg	103			84	121	
	n-Butylbenzene	20	21.2	ug/Kg	106			78	126	
	1,2-Dichlorobenzene	20	19.8	ug/Kg	99			83	124	
	1,2-Dibromo-3-Chloropropane	20	19.7	ug/Kg	99			66	134	
	1,2,4-Trichlorobenzene	20	19.3	ug/Kg	97			78	127	
	1,2,3-Trichlorobenzene	20	19.2	ug/Kg	96			70	137	