



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LABE01
Lab Code: CHEM Case No.: 03645 SAS No.: 03645 SDG NO.: 03645
Lab File ID: VD076766.D Date Analyzed: 07/18/2023
Instrument ID: MSVOA_D Time Analyzed: 09:13
GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	206994	7.88	357390	8.78	317730	11.58
UPPER LIMIT	413988	8.375	714780	9.275	635460	12.081
LOWER LIMIT	103497	7.375	178695	8.275	158865	11.081
EPA SAMPLE NO.						
DUP	147654	7.88	354882	8.78	328863	11.58
VD0718SBL01	163878	7.88	390285	8.78	373979	11.58
VD0718SBS01	187561	7.88	339123	8.78	301791	11.58

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.