



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LABE01
Lab Code: CHEM Case No.: 03645 SAS No.: 03645 SDG NO.: 03645
Lab File ID: VN078550.D Date Analyzed: 07/18/2023
Instrument ID: MSVOA_N Time Analyzed: 10:29
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	483266	8.24	784393	9.11	713959	11.87
UPPER LIMIT	966532	8.735	1568790	9.606	1427920	12.37
LOWER LIMIT	241633	7.735	392197	8.606	356980	11.37
EPA SAMPLE NO.						
RINSATE-BLANK	401033	8.24	752774	9.11	693621	11.87
VN0718WBL01	402785	8.24	700602	9.11	679555	11.87
VN0718WBS02	432335	8.23	742826	9.11	745463	11.87
VN0718WBSD02	485193	8.24	833352	9.11	717025	11.87

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.