



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LABE01
Lab Code: CHEM Case No.: 03645 SAS No.: 03645 SDG NO.: 03645
Lab File ID: VD076766.D Date Analyzed: 07/18/2023
Instrument ID: MSVOA_D Time Analyzed: 09:13
GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	151547	13.516				
UPPER LIMIT	303094	14.016				
LOWER LIMIT	75773.5	13.016				
EPA SAMPLE NO.						
DUP	141705	13.52				
VD0718SBL01	159933	13.52				
VD0718SBS01	145517	13.52				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.