



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LABE01
Lab Code: CHEM Case No.: 03645 SAS No.: 03645 SDG NO.: 03645
Lab File ID: VN078550.D Date Analyzed: 07/18/2023
Instrument ID: MSVOA_N Time Analyzed: 10:29
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	297857	13.8				
UPPER LIMIT	595714	14.3				
LOWER LIMIT	148929	13.3				
EPA SAMPLE NO.						
RINSATE-BLANK	242621	13.80				
VN0718WBL01	243985	13.79				
VN0718WBS02	251794	13.79				
VN0718WBSD02	285080	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.