



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: LABE01
Lab Code: CHEM Case No.: 03645 SAS No.: 03645 SDG NO.: 03645
Lab File ID: VN078448.D BFB Injection Date: 07/10/2023
Instrument ID: MSVOA_N BFB Injection Time: 11:48
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.6
75	30.0 - 60.0% of mass 95	54.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.2
173	Less than 2.0% of mass 174	0.6 (0.8) 1
174	50.0 - 100.0% of mass 95	77.9
175	5.0 - 9.0% of mass 174	6.1 (7.8) 1
176	95.0 - 101.0% of mass 174	74.2 (95.3) 1
177	5.0 - 9.0% of mass 176	4.8 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC100	VSTDICC100	VN078449.D	07/10/2023	12:21
VSTDICCC050	VSTDICCC050	VN078450.D	07/10/2023	12:46
VSTDICC040	VSTDICC040	VN078451.D	07/10/2023	13:10
VSTDICC020	VSTDICC020	VN078452.D	07/10/2023	13:34
VSTDICC005	VSTDICC005	VN078453.D	07/10/2023	13:58
VSTDICC001	VSTDICC001	VN078454.D	07/10/2023	14:46