



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LABE01
Lab Code: CHEM Case No.: 03645 SAS No.: 03645 SDG NO.: 03645
Lab File ID: VD076745.D Date Analyzed: 07/17/2023
Instrument ID: MSVOA_D Time Analyzed: 10:19
GC Column: RTX-VMS ID: 0.18 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	160211	13.516				
UPPER LIMIT	320422	14.016				
LOWER LIMIT	80105.5	13.016				
EPA SAMPLE NO.						
SB-02-(3-5)	187559	13.52				
SB-04-(1-5)	170608	13.52				
SB-07-(1-3)	128769	13.52				
SB-08-(0.5-2.0)	92343	13.52				
SB-10-(0.5-2.0)	139423	13.52				
SB-04-(1-5)MS	102048	13.52				
SB-04-(1-5)MSD	117737	13.52				
VD0717SBL01	184069	13.52				
VD0717SBS01	159296	13.52				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.