



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: O3649 SAS No.: O3649 SDG NO.: O3649

Continuing Calib Date: 07/18/2023 Initial Calibration Date(s): 07/01/2023 07/01/2023

Continuing Calib Time: 11:15 Initial Calibration Time(s): 03:04 09:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.50	4.50	4.40	4.60	0.00
Aroclor-1248-2 (2)	4.77	4.77	4.67	4.87	0.00
Aroclor-1248-3 (3)	4.96	4.96	4.86	5.06	0.00
Aroclor-1248-4 (4)	5.35	5.35	5.25	5.45	0.00
Aroclor-1248-5 (5)	5.39	5.39	5.29	5.49	0.00
Tetrachloro-m-xylene	3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl	8.59	8.58	8.48	8.68	-0.01



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: O3649 SAS No.: O3649 SDG NO.: O3649

Continuing Calib Date: 07/18/2023 Initial Calibration Date(s): 07/01/2023 07/01/2023

Continuing Calib Time: 11:15 Initial Calibration Time(s): 03:04 09:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	3.76	3.76	3.66	3.86	0.00
Aroclor-1248-2 (2)	3.98	3.99	3.89	4.09	0.01
Aroclor-1248-3 (3)	4.02	4.02	3.92	4.12	0.00
Aroclor-1248-4 (4)	4.18	4.18	4.08	4.28	0.00
Aroclor-1248-5 (5)	4.55	4.55	4.45	4.65	0.00
Tetrachloro-m-xylene	2.76	2.76	2.66	2.86	0.00
Decachlorobiphenyl	7.53	7.53	7.43	7.63	0.00