



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: O3649 SAS No.: O3649 SDG NO.: O3649

Continuing Calib Date: 07/18/2023 Initial Calibration Date(s): 07/01/2023 07/01/2023

Continuing Calib Time: 22:07 Initial Calibration Time(s): 03:04 09:39

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.32       | 5.33      | 5.23      | 5.43 | 0.01       |
| Aroclor-1254-2 (2)   | 5.54       | 5.54      | 5.44      | 5.64 | 0.00       |
| Aroclor-1254-3 (3)   | 5.90       | 5.90      | 5.80      | 6.00 | 0.01       |
| Aroclor-1254-4 (4)   | 6.18       | 6.18      | 6.08      | 6.28 | 0.00       |
| Aroclor-1254-5 (5)   | 6.59       | 6.59      | 6.49      | 6.69 | 0.00       |
| Tetrachloro-m-xylene | 3.40       | 3.40      | 3.30      | 3.50 | 0.00       |
| Decachlorobiphenyl   | 8.58       | 8.58      | 8.48      | 8.68 | 0.00       |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: O3649 SAS No.: O3649 SDG NO.: O3649

Continuing Calib Date: 07/18/2023 Initial Calibration Date(s): 07/01/2023 07/01/2023

Continuing Calib Time: 22:07 Initial Calibration Time(s): 03:04 09:39

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 4.51       | 4.52      | 4.42      | 4.62 | 0.01       |
| Aroclor-1254-2 (2)   | 4.66       | 4.66      | 4.56      | 4.76 | 0.00       |
| Aroclor-1254-3 (3)   | 5.04       | 5.04      | 4.94      | 5.14 | 0.00       |
| Aroclor-1254-4 (4)   | 5.27       | 5.27      | 5.17      | 5.37 | 0.00       |
| Aroclor-1254-5 (5)   | 5.67       | 5.68      | 5.58      | 5.78 | 0.01       |
| Tetrachloro-m-xylene | 2.75       | 2.76      | 2.66      | 2.86 | 0.01       |
| Decachlorobiphenyl   | 7.52       | 7.53      | 7.43      | 7.63 | 0.01       |