



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: O3649 SAS No.: O3649 SDG NO.: O3649

Continuing Calib Date: 07/19/2023 Initial Calibration Date(s): 07/01/2023 07/01/2023

Continuing Calib Time: 01:17 Initial Calibration Time(s): 03:04 09:39

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1254-1 (1)	5.33	5.33	5.23	5.43	0.01
Aroclor-1254-2 (2)	5.54	5.54	5.44	5.64	0.00
Aroclor-1254-3 (3)	5.90	5.90	5.80	6.00	0.01
Aroclor-1254-4 (4)	6.18	6.18	6.08	6.28	0.00
Aroclor-1254-5 (5)	6.59	6.59	6.49	6.69	0.00
Tetrachloro-m-xylene	3.40	3.40	3.30	3.50	0.00
Decachlorobiphenyl	8.58	8.58	8.48	8.68	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: O3649 SAS No.: O3649 SDG NO.: O3649

Continuing Calib Date: 07/19/2023 Initial Calibration Date(s): 07/01/2023 07/01/2023

Continuing Calib Time: 01:17 Initial Calibration Time(s): 03:04 09:39

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	4.51	4.52	4.42	4.62	0.01
Aroclor-1254-2 (2)	4.66	4.66	4.56	4.76	0.00
Aroclor-1254-3 (3)	5.04	5.04	4.94	5.14	0.00
Aroclor-1254-4 (4)	5.27	5.27	5.17	5.37	0.00
Aroclor-1254-5 (5)	5.67	5.68	5.58	5.78	0.01
Tetrachloro-m-xylene	2.75	2.76	2.66	2.86	0.01
Decachlorobiphenyl	7.52	7.53	7.43	7.63	0.01