



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: O3649 SAS No.: O3649 SDG NO.: O3649

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 07/01/2023 07/01/2023

Client Sample No.: CCAL02 Date Analyzed: 07/18/2023

Lab Sample No.: AR1660CCC500 Data File : PQ062344.D Time Analyzed: 15:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.502	4.404	4.604	455.760	500.000	-8.8
Aroclor-1016-2	4.521	4.424	4.624	449.340	500.000	-10.1
Aroclor-1016-3	4.579	4.481	4.681	451.930	500.000	-9.6
Aroclor-1016-4	4.671	4.573	4.773	454.100	500.000	-9.2
Aroclor-1016-5	4.958	4.860	5.060	464.250	500.000	-7.2
Aroclor-1260-1	6.060	5.963	6.163	459.120	500.000	-8.2
Aroclor-1260-2	6.321	6.223	6.423	456.430	500.000	-8.7
Aroclor-1260-3	6.674	6.576	6.776	465.460	500.000	-6.9
Aroclor-1260-4	6.891	6.793	6.993	469.240	500.000	-6.2
Aroclor-1260-5	7.203	7.104	7.304	472.820	500.000	-5.4
Decachlorobiphenyl	8.582	8.484	8.684	45.700	50.000	-8.6
Tetrachloro-m-xylene	3.402	3.304	3.504	44.040	50.000	-11.9



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: O3649 SAS No.: O3649 SDG NO.: O3649

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/01/2023 07/01/2023

Client Sample No.: CCAL02 Date Analyzed: 07/18/2023

Lab Sample No.: AR1660CCC500 Data File : PQ062344.D Time Analyzed: 15:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.757	3.661	3.861	474.100	500.000	-5.2
Aroclor-1016-2	3.772	3.676	3.876	482.260	500.000	-3.5
Aroclor-1016-3	3.932	3.836	4.036	474.720	500.000	-5.1
Aroclor-1016-4	3.982	3.886	4.086	471.680	500.000	-5.7
Aroclor-1016-5	4.176	4.080	4.280	466.770	500.000	-6.6
Aroclor-1260-1	5.169	5.073	5.273	484.360	500.000	-3.1
Aroclor-1260-2	5.360	5.263	5.463	476.990	500.000	-4.6
Aroclor-1260-3	5.500	5.404	5.604	489.360	500.000	-2.1
Aroclor-1260-4	5.962	5.867	6.067	498.090	500.000	-0.4
Aroclor-1260-5	6.209	6.113	6.313	501.990	500.000	0.4
Decachlorobiphenyl	7.525	7.429	7.629	48.730	50.000	-2.5
Tetrachloro-m-xylene	2.756	2.659	2.859	49.150	50.000	-1.7