



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: O3649 SAS No.: O3649 SDG NO.: O3649

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 07/01/2023 07/01/2023

Client Sample No.: CCAL03 Date Analyzed: 07/18/2023

Lab Sample No.: AR1660CCC500 Data File : PQ062366.D Time Analyzed: 21:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.502	4.404	4.604	468.090	500.000	-6.4
Aroclor-1016-2	4.521	4.424	4.624	462.940	500.000	-7.4
Aroclor-1016-3	4.579	4.481	4.681	461.900	500.000	-7.6
Aroclor-1016-4	4.671	4.573	4.773	467.300	500.000	-6.5
Aroclor-1016-5	4.958	4.860	5.060	470.260	500.000	-5.9
Aroclor-1260-1	6.061	5.963	6.163	469.520	500.000	-6.1
Aroclor-1260-2	6.321	6.223	6.423	468.840	500.000	-6.2
Aroclor-1260-3	6.675	6.576	6.776	472.030	500.000	-5.6
Aroclor-1260-4	6.891	6.793	6.993	474.710	500.000	-5.1
Aroclor-1260-5	7.203	7.104	7.304	486.590	500.000	-2.7
Decachlorobiphenyl	8.581	8.484	8.684	48.760	50.000	-2.5
Tetrachloro-m-xylene	3.401	3.304	3.504	45.410	50.000	-9.2



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: O3649 SAS No.: O3649 SDG NO.: O3649

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/01/2023 07/01/2023

Client Sample No.: CCAL03 Date Analyzed: 07/18/2023

Lab Sample No.: AR1660CCC500 Data File : PQ062366.D Time Analyzed: 21:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.756	3.661	3.861	478.180	500.000	-4.4
Aroclor-1016-2	3.772	3.676	3.876	478.850	500.000	-4.2
Aroclor-1016-3	3.931	3.836	4.036	472.580	500.000	-5.5
Aroclor-1016-4	3.981	3.886	4.086	469.160	500.000	-6.2
Aroclor-1016-5	4.175	4.080	4.280	478.220	500.000	-4.4
Aroclor-1260-1	5.169	5.073	5.273	499.800	500.000	0.0
Aroclor-1260-2	5.359	5.263	5.463	495.400	500.000	-0.9
Aroclor-1260-3	5.500	5.404	5.604	500.630	500.000	0.1
Aroclor-1260-4	5.961	5.867	6.067	507.020	500.000	1.4
Aroclor-1260-5	6.208	6.113	6.313	515.780	500.000	3.2
Decachlorobiphenyl	7.524	7.429	7.629	50.960	50.000	1.9
Tetrachloro-m-xylene	2.755	2.659	2.859	50.060	50.000	0.1