

Surrogate Summary

SDG No.: **O3649**

Client: **ATC Group Services LLC**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PQ061834.D	PIBLK-PQ061834.D	Tetrachloro-m-xylene	1	20	20.0	100		60	140
		Decachlorobiphenyl	1	20	20.9	105		60	140
		Tetrachloro-m-xylene	2	20	20.3	101		60	140
		Decachlorobiphenyl	2	20	20.5	102		60	140
I.BLK-PQ062334.D	PIBLK-PQ062334.D	Tetrachloro-m-xylene	1	20	15.0	75		60	140
		Decachlorobiphenyl	1	20	17.2	86		60	140
		Tetrachloro-m-xylene	2	20	17.7	89		60	140
		Decachlorobiphenyl	2	20	18.5	92		60	140
PB154254BL	PB154254BL	Tetrachloro-m-xylene	1	20	17.0	85		40	162
		Decachlorobiphenyl	1	20	19.4	97		32	176
		Tetrachloro-m-xylene	2	20	19.1	96		40	162
		Decachlorobiphenyl	2	20	20.4	102		32	176
PB154254BS	PB154254BS	Tetrachloro-m-xylene	1	20	16.9	84		40	162
		Decachlorobiphenyl	1	20	19.0	95		32	176
		Tetrachloro-m-xylene	2	20	18.2	91		40	162
		Decachlorobiphenyl	2	20	19.8	99		32	176
I.BLK-PQ062346.D	PIBLK-PQ062346.D	Tetrachloro-m-xylene	1	20	15.9	79		60	140
		Decachlorobiphenyl	1	20	17.1	85		60	140
		Tetrachloro-m-xylene	2	20	18.4	92		60	140
		Decachlorobiphenyl	2	20	18.0	90		60	140
O3645-09MS	SB-04-(1-5)MS	Tetrachloro-m-xylene	1	20	18.7	93		40	162
		Decachlorobiphenyl	1	20	18.7	93		32	176
		Tetrachloro-m-xylene	2	20	20.3	102		40	162
		Decachlorobiphenyl	2	20	19.7	98		32	176
O3645-10MSD	SB-04-(1-5)MSD	Tetrachloro-m-xylene	1	20	18.9	94		40	162
		Decachlorobiphenyl	1	20	19.0	95		32	176
		Tetrachloro-m-xylene	2	20	20.3	101		40	162
		Decachlorobiphenyl	2	20	19.7	99		32	176
O3649-01	X182-S5	Tetrachloro-m-xylene	1	20	16.2	81		40	162
		Decachlorobiphenyl	1	20	14.1	71		32	176
		Tetrachloro-m-xylene	2	20	18.3	91		40	162
		Decachlorobiphenyl	2	20	16.0	80		32	176
O3649-02	X182-B4	Tetrachloro-m-xylene	1	20	15.5	77		40	162
		Decachlorobiphenyl	1	20	10.8	54		32	176
		Tetrachloro-m-xylene	2	20	17.2	86		40	162
		Decachlorobiphenyl	2	20	12.5	63		32	176
O3649-03	X182-S3	Tetrachloro-m-xylene	1	20	17.1	85		40	162
		Decachlorobiphenyl	1	20	18.4	92		32	176
		Tetrachloro-m-xylene	2	20	19.0	95		40	162
		Decachlorobiphenyl	2	20	18.6	93		32	176
O3649-04	X182-B3	Tetrachloro-m-xylene	1	20	17.3	87		40	162
		Decachlorobiphenyl	1	20	17.4	87		32	176
		Tetrachloro-m-xylene	2	20	19.4	97		40	162



Surrogate Summary

SDG No.: O3649

Client: ATC Group Services LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
O3649-04	X182-B3	Decachlorobiphenyl	2	20	19.3	97		32	176
O3649-05	X182-B3-DUP	Tetrachloro-m-xylene	1	20	24.6	123		40	162
		Decachlorobiphenyl	1	20	29.6	148		32	176
		Tetrachloro-m-xylene	2	20	27.3	137		40	162
		Decachlorobiphenyl	2	20	32.3	162		32	176
O3649-06	X182-S4	Tetrachloro-m-xylene	1	20	17.0	85		40	162
		Decachlorobiphenyl	1	20	15.8	79		32	176
		Tetrachloro-m-xylene	2	20	19.1	95		40	162
		Decachlorobiphenyl	2	20	17.4	87		32	176
I.BLK-PQ062368.D	PIBLK-PQ062368.D	Tetrachloro-m-xylene	1	20	16.7	84		60	140
		Decachlorobiphenyl	1	20	18.3	92		60	140
		Tetrachloro-m-xylene	2	20	19.3	96		60	140
		Decachlorobiphenyl	2	20	19.7	98		60	140