



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG No.: O4699

Calibration Sequence : FE100323		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	151113759	151114	FE044569.D
500	77467580	154935	FE044570.D
200	31530013	157650	FE044571.D
100	17278740	172787	FE044572.D
50	8339805	166796	FE044573.D
AVG RF : 160656		% RSD : 5.547	AVG RT : 15.0158