



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG No.: O4699

Calibration Sequence : FG092123		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	110623730	110624	FG012937.D
500	56177632	112355	FG012938.D
200	23454105	117271	FG012939.D
100	12899738	128997	FG012940.D
50	6301779	126036	FG012941.D
AVG RF : 119057		% RSD : 6.859	AVG RT : 15.078