



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC100523AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC064737.D

Aliphatic C9-C12	11597745.000	60.000	3.222	6.489	193295.750	179334.669	-7.785
Aliphatic C12-C16	7552626.000	40.000	6.490	9.871	188815.650	178597.012	-5.722
Aliphatic C16-C21	10711290.000	60.000	9.872	13.217	178521.500	172340.408	-3.587
Aliphatic C21-C28	12884729.000	80.000	13.218	16.860	161059.113	155578.573	-3.523
Aliphatic C28-C40	17271227.000	120.000	16.861	21.688	143926.892	141245.657	-1.898
Aliphatic EPH	60017617.000	360.000	3.222	21.688	166715.603	160111.527	-4.125