



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC100523AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC064746.D

Aliphatic C9-C12	11708115.000	60.000	3.222	6.489	195135.250	179334.669	-8.811
Aliphatic C12-C16	7691249.000	40.000	6.490	9.871	192281.225	178597.012	-7.662
Aliphatic C16-C21	11000117.000	60.000	9.872	13.217	183335.283	172340.408	-6.380
Aliphatic C21-C28	13299186.000	80.000	13.218	16.860	166239.825	155578.573	-6.853
Aliphatic C28-C40	17768352.000	120.000	16.861	21.688	148069.600	141245.657	-4.831
Aliphatic EPH	61467019.000	360.000	3.222	21.688	170741.719	160111.527	-6.639