



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FC100523AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC064753.D

Aliphatic C9-C12	11708221.000	60.000	3.222	6.489	195137.017	179334.669	-8.812
Aliphatic C12-C16	7684468.000	40.000	6.490	9.871	192111.700	178597.012	-7.567
Aliphatic C16-C21	11000792.000	60.000	9.872	13.217	183346.533	172340.408	-6.386
Aliphatic C21-C28	13342557.000	80.000	13.218	16.860	166781.963	155578.573	-7.201
Aliphatic C28-C40	17820141.000	120.000	16.861	21.688	148501.175	141245.657	-5.137
Aliphatic EPH	61556179.000	360.000	3.222	21.688	170989.386	160111.527	-6.794