



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FD100523AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD046652.D

Aromatic C10-C12	8291586.000	40.000	4.202	5.928	207289.650	188965.877	-9.697
Aromatic C12-C16	12311828.000	60.000	5.929	8.541	205197.133	186001.899	-10.320
Aromatic C16-C21	14965721.000	80.000	8.542	12.804	187071.513	168464.674	-11.045
Aromatic C21-C36	27680237.000	180.000	12.805	18.192	153779.094	140182.042	-9.700
Aromatic EPH	63249372.000	360.000	4.202	18.192	175692.700	159524.140	-10.135