



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FD100523AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD046659.D

Aromatic C10-C12	7608608.000	40.000	4.202	5.928	190215.200	188965.877	-0.661
Aromatic C12-C16	11279647.000	60.000	5.929	8.541	187994.117	186001.899	-1.071
Aromatic C16-C21	13801181.000	80.000	8.542	12.804	172514.763	168464.674	-2.404
Aromatic C21-C36	25675415.000	180.000	12.805	18.192	142641.194	140182.042	-1.754
Aromatic EPH	58364851.000	360.000	4.202	18.192	162124.586	159524.140	-1.630