



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FD100523AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD046664.D

Aromatic C10-C12	7528197.000	40.000	4.202	5.928	188204.925	188965.877	0.403
Aromatic C12-C16	11200163.000	60.000	5.929	8.541	186669.383	186001.899	-0.359
Aromatic C16-C21	13650383.000	80.000	8.542	12.804	170629.788	168464.674	-1.285
Aromatic C21-C36	25508232.000	180.000	12.805	18.192	141712.400	140182.042	-1.092
Aromatic EPH	57886975.000	360.000	4.202	18.192	160797.153	159524.140	-0.798