



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE100623AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE044649.D

Aliphatic C9-C12	9520127.000	60.000	3.204	6.824	158668.783	164757.986	3.696
Aliphatic C12-C16	6491415.000	40.000	6.825	10.267	162285.375	166787.277	2.699
Aliphatic C16-C21	9502002.000	60.000	10.268	13.643	158366.700	162421.243	2.496
Aliphatic C21-C28	11790794.000	80.000	13.644	17.315	147384.925	148979.105	1.070
Aliphatic C28-C40	18319586.000	120.000	17.316	22.263	152663.217	149177.608	-2.337
Aliphatic EPH	55623924.000	360.000	3.204	22.263	154510.900	155894.128	0.887