



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Continuing Calibration Report for SequenceID : FE100623AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE044671.D

Aliphatic C9-C12	9114735.000	60.000	3.204	6.824	151912.250	164757.986	7.797
Aliphatic C12-C16	6245729.000	40.000	6.825	10.267	156143.225	166787.277	6.382
Aliphatic C16-C21	9153881.000	60.000	10.268	13.643	152564.683	162421.243	6.069
Aliphatic C21-C28	11526235.000	80.000	13.644	17.315	144077.938	148979.105	3.290
Aliphatic C28-C40	19114431.000	120.000	17.316	22.263	159286.925	149177.608	-6.777
Aliphatic EPH	55155011.000	360.000	3.204	22.263	153208.364	155894.128	1.723