



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/09/2023 Initial Calibration Date(s): 09/27/2023 09/27/2023

Continuing Calib Time: 10:21 Initial Calibration Time(s): 12:17 19:52

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1254-1 (1)   | 6.52       | 6.52      | 6.42      | 6.62  | 0.00       |
| Aroclor-1254-2 (2)   | 6.74       | 6.74      | 6.64      | 6.84  | 0.00       |
| Aroclor-1254-3 (3)   | 7.10       | 7.11      | 7.01      | 7.21  | 0.01       |
| Aroclor-1254-4 (4)   | 7.39       | 7.40      | 7.30      | 7.50  | 0.01       |
| Aroclor-1254-5 (5)   | 7.81       | 7.82      | 7.72      | 7.92  | 0.01       |
| Tetrachloro-m-xylene | 4.48       | 4.49      | 4.39      | 4.59  | 0.01       |
| Decachlorobiphenyl   | 10.30      | 10.31     | 10.21     | 10.41 | 0.01       |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/09/2023 Initial Calibration Date(s): 09/27/2023 09/27/2023

Continuing Calib Time: 10:21 Initial Calibration Time(s): 12:17 19:52

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.52       | 5.53      | 5.43      | 5.63 | 0.01       |
| Aroclor-1254-2 (2)   | 5.67       | 5.67      | 5.57      | 5.77 | 0.00       |
| Aroclor-1254-3 (3)   | 6.07       | 6.08      | 5.98      | 6.18 | 0.01       |
| Aroclor-1254-4 (4)   | 6.30       | 6.31      | 6.21      | 6.41 | 0.01       |
| Aroclor-1254-5 (5)   | 6.72       | 6.73      | 6.63      | 6.83 | 0.01       |
| Tetrachloro-m-xylene | 3.64       | 3.64      | 3.54      | 3.74 | 0.01       |
| Decachlorobiphenyl   | 8.64       | 8.65      | 8.55      | 8.75 | 0.01       |