



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/10/2023 Initial Calibration Date(s): 09/27/2023 09/27/2023

Continuing Calib Time: 02:56 Initial Calibration Time(s): 12:17 19:52

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|-------|------------|
| Aroclor-1016-1       | (1) | 5.66       | 5.66      | 5.56                   | 5.76  | 0.00       |
| Aroclor-1016-2       | (2) | 5.68       | 5.68      | 5.58                   | 5.78  | 0.00       |
| Aroclor-1016-3       | (3) | 5.74       | 5.74      | 5.64                   | 5.84  | 0.00       |
| Aroclor-1016-4       | (4) | 5.84       | 5.84      | 5.74                   | 5.94  | 0.00       |
| Aroclor-1016-5       | (5) | 6.14       | 6.14      | 6.04                   | 6.24  | 0.00       |
| Aroclor-1260-1       | (1) | 7.27       | 7.27      | 7.17                   | 7.37  | 0.00       |
| Aroclor-1260-2       | (2) | 7.52       | 7.53      | 7.43                   | 7.63  | 0.01       |
| Aroclor-1260-3       | (3) | 7.88       | 7.89      | 7.79                   | 7.99  | 0.01       |
| Aroclor-1260-4       | (4) | 8.11       | 8.11      | 8.01                   | 8.21  | 0.00       |
| Aroclor-1260-5       | (5) | 8.44       | 8.44      | 8.34                   | 8.54  | 0.00       |
| Tetrachloro-m-xylene |     | 4.48       | 4.49      | 4.39                   | 4.59  | 0.01       |
| Decachlorobiphenyl   |     | 10.30      | 10.29     | 10.19                  | 10.39 | -0.01      |



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/10/2023 Initial Calibration Date(s): 09/27/2023 09/27/2023

Continuing Calib Time: 02:56 Initial Calibration Time(s): 12:17 19:52

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.72       | 4.72      | 4.62      | 4.82 | 0.00       |
| Aroclor-1016-2       | (2) | 4.74       | 4.74      | 4.64      | 4.84 | 0.00       |
| Aroclor-1016-3       | (3) | 4.91       | 4.92      | 4.82      | 5.02 | 0.01       |
| Aroclor-1016-4       | (4) | 4.95       | 4.96      | 4.86      | 5.06 | 0.01       |
| Aroclor-1016-5       | (5) | 5.17       | 5.17      | 5.07      | 5.27 | 0.00       |
| Aroclor-1260-1       | (1) | 6.20       | 6.21      | 6.11      | 6.31 | 0.01       |
| Aroclor-1260-2       | (2) | 6.39       | 6.40      | 6.30      | 6.50 | 0.01       |
| Aroclor-1260-3       | (3) | 6.54       | 6.55      | 6.45      | 6.65 | 0.01       |
| Aroclor-1260-4       | (4) | 7.02       | 7.02      | 6.92      | 7.12 | 0.00       |
| Aroclor-1260-5       | (5) | 7.26       | 7.27      | 7.17      | 7.37 | 0.01       |
| Tetrachloro-m-xylene |     | 3.63       | 3.64      | 3.54      | 3.74 | 0.01       |
| Decachlorobiphenyl   |     | 8.64       | 8.65      | 8.55      | 8.75 | 0.01       |