



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/10/2023 Initial Calibration Date(s): 09/27/2023 09/27/2023

Continuing Calib Time: 19:20 Initial Calibration Time(s): 12:17 19:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.65	5.66	5.56	5.76	0.01
Aroclor-1016-2	(2)	5.68	5.68	5.58	5.78	0.00
Aroclor-1016-3	(3)	5.74	5.74	5.64	5.84	0.00
Aroclor-1016-4	(4)	5.84	5.84	5.74	5.94	0.00
Aroclor-1016-5	(5)	6.13	6.14	6.04	6.24	0.01
Aroclor-1260-1	(1)	7.26	7.27	7.17	7.37	0.01
Aroclor-1260-2	(2)	7.52	7.53	7.43	7.63	0.01
Aroclor-1260-3	(3)	7.88	7.89	7.79	7.99	0.01
Aroclor-1260-4	(4)	8.11	8.11	8.01	8.21	0.00
Aroclor-1260-5	(5)	8.44	8.44	8.34	8.54	0.00
Tetrachloro-m-xylene		4.48	4.49	4.39	4.59	0.01
Decachlorobiphenyl		10.29	10.29	10.19	10.39	0.00



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: O4699 SAS No.: O4699 SDG NO.: O4699

Continuing Calib Date: 10/10/2023 Initial Calibration Date(s): 09/27/2023 09/27/2023

Continuing Calib Time: 19:20 Initial Calibration Time(s): 12:17 19:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1016-1 (1)	4.72	4.72	4.62	4.82	0.00
Aroclor-1016-2 (2)	4.74	4.74	4.64	4.84	0.00
Aroclor-1016-3 (3)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-4 (4)	4.95	4.96	4.86	5.06	0.01
Aroclor-1016-5 (5)	5.17	5.17	5.07	5.27	0.00
Aroclor-1260-1 (1)	6.20	6.21	6.11	6.31	0.01
Aroclor-1260-2 (2)	6.39	6.40	6.30	6.50	0.01
Aroclor-1260-3 (3)	6.54	6.55	6.45	6.65	0.01
Aroclor-1260-4 (4)	7.02	7.02	6.92	7.12	0.00
Aroclor-1260-5 (5)	7.26	7.27	7.17	7.37	0.01
Tetrachloro-m-xylene	3.63	3.64	3.54	3.74	0.01
Decachlorobiphenyl	8.64	8.65	8.55	8.75	0.01